

LB126131

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

63297

A= 0.0000e+000

B= 1.5853e-004

C= -1.2105e-002

Rho= 0.9997053

Accept=Accepted

Max. Conc.: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	-0.006	-0.006	40	0.000	40					
0.2	0.200	0.237	0.037	1574	0.0 %	1574					19
2.5	2.500	2.571	0.071	16292	0.0 %	16292					3
5.0	5.000	4.808	-0.192	30407	0.0 %	30407					-4
7.5	7.500	7.567	0.067	47808	0.0 %	47808					1
10.0	10.000	10.022	0.022	63297	0.0 %	63297					0

LB126135 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jun 2023 20:05:34

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	144	-	0.0000	7470A	PPB	22 Jun 2023 20:08:29	S	Std
0.2 - 1	50.2	711	-	0.2000	7470A	PPB	22 Jun 2023 20:10:45	S	Std
2.5 - 1	52.5	12400	-	2.5000	7470A	PPB	22 Jun 2023 20:13:02	S	Std
5.0 - 1	55	23816	-	5.0000	7470A	PPB	22 Jun 2023 20:15:19	S	Std
7.5 - 1	57.5	35552	-	7.5000	7470A	PPB	22 Jun 2023 20:17:35	S	Std
10.0 - 1	50	47983	-	10.0000	7470A	PPB	22 Jun 2023 20:19:52	S	Std
ICV37 - 1	ICV37	18118	3.7848	-	7470A	PPB	22 Jun 2023 20:22:40	U	SMPL
ICB37 - 1	ICB37	37	-0.0006	-	7470A	PPB	22 Jun 2023 20:24:55	U	SMPL
CCV14 - 1	CCV14	23526	4.9171	-	7470A	PPB	22 Jun 2023 20:27:11	U	SMPL
CCB14 - 1	CCB14	60	0.0042	-	7470A	PPB	22 Jun 2023 20:29:25	U	SMPL
CRA - 1	CRA	856	0.1708	-	7470A	PPB	22 Jun 2023 20:31:41	U	SMPL
HighStd - 1	HighStd	47490	9.9342	-	7470A	PPB	22 Jun 2023 20:33:56	U	SMPL
ChkStd - 1	ChkStd	32630	6.8231	-	7470A	PPB	22 Jun 2023 20:36:11	U	SMPL
PB153706BL - 1	PBW	-10	-0.0105	-	7470A	PPB	22 Jun 2023 20:38:29	U	SMPL
PB153706BS - 1	LCSW	16484	3.4427	-	7470A	PPB	22 Jun 2023 20:40:46	U	SMPL
O3304-04 - 1	JC-WOOD-0616	1340	0.2722	-	7470A	PPB	22 Jun 2023 20:43:02	U	SMPL
O3304-04DUP - 1	JC-WOOD-0616DUP	1416	0.2881	-	7470A	PPB	22 Jun 2023 20:45:19	U	SMPL
O3304-04MS - 1	JC-WOOD-0616MS	18519	3.8688	-	7470A	PPB	22 Jun 2023 20:47:36	U	SMPL
O3304-04MSD - 1	JC-WOOD-0616MSD	18278	3.8183	-	7470A	PPB	22 Jun 2023 20:49:52	U	SMPL
CCV15 - 1	CCV15	22888	4.7835	-	7470A	PPB	22 Jun 2023 20:52:08	U	SMPL
CCB15 - 1	CCB15	32	-0.0017	-	7470A	PPB	22 Jun 2023 20:54:25	U	SMPL
O3304-04LX5 - 1		317	0.0580	-	7470A	PPB	22 Jun 2023 20:56:42	U	SMPL
O3304-04A - 1		19079	3.9860	-	7470A	PPB	22 Jun 2023 20:58:57	U	SMPL
PB153530TB - 1	PB153530TB	22	-0.0038	-	7470A	PPB	22 Jun 2023 21:06:51	U	SMPL
CCV15 - 1	CCV15	23667	4.9466	-	7470A	PPB	22 Jun 2023 21:09:06	U	SMPL
CCB16 - 1	CCB16	-89	-0.0270	-	7470A	PPB	22 Jun 2023 21:11:21	U	SMPL