



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126156

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/23/2023	14:25
2	CAL2	1	Water	NA	NA	20.7	7.00	06/23/2023	14:26
3	CAL3	1	Water	NA	NA	20.7	10.02	06/23/2023	14:27
4	ICV	1	Water	NA	NA	20.7	7.02	06/23/2023	14:30
5	CCV1	1	Water	NA	NA	20.7	2.01	06/23/2023	14:33
6	O3381-01	1	Solid	20.02	20	22.8	7.54	06/23/2023	14:35
7	O3381-01DUP	1	Solid	20.03	20	22.9	7.55	06/23/2023	14:36
8	O3381-02	1	Solid	20.02	20	22.8	7.28	06/23/2023	14:40
9	O3381-03	1	Solid	20.03	20	22.6	7.37	06/23/2023	14:56
10	CCV2	1	Water	NA	NA	20.6	12.02	06/23/2023	15:00

VP126156

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o3393 WorkList ID : 171270 Department : Wet-Chemistry Date : 06-23-2023 14:15:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3381-01	COMP1-0622	Solid	Corrosivity	Cool 4 deg C	PSEG03	H11	06/22/2023	9045D
O3381-02	COMP2-0622	Solid	Corrosivity	Cool 4 deg C	PSEG03	H11	06/22/2023	9045D
O3381-03	8352	Solid	Corrosivity	Cool 4 deg C	PSEG03	H11	06/22/2023	9045D

Date/Time 06123123 14:20
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC

Date/Time 06123123 16:00
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC