



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB126157

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/23/2023	14:25
2	CAL2	1	Water	NA	NA	20.7	7.00	06/23/2023	14:26
3	CAL3	1	Water	NA	NA	20.7	10.02	06/23/2023	14:27
4	ICV	1	Water	NA	NA	20.7	7.02	06/23/2023	14:30
5	CCV1	1	Water	NA	NA	20.9	2.02	06/23/2023	15:10
6	O3381-04	1	Water	NA	NA	23.1	7.54	06/23/2023	15:15
7	O3381-04DUP	1	Water	NA	NA	23.2	7.55	06/23/2023	15:16
8	CCV2	1	Water	NA	NA	20.7	12.02	06/23/2023	15:20

VB126154

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o3381 w WorkList ID : 171271 Department : Wet-Chemistry Date : 06-23-2023 14:16:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3381-04	8327	Water	Corrosivity	Cool 4 deg C	PSEG03	H11	06/22/2023	9040C

VB126

06/23/23

Date/Time

Raw Sample Received by: JD/CDC

Raw Sample Relinquished by: F.C (sm)

VB.00

06/23/23

Date/Time

Raw Sample Received by: F.C (sm)

Raw Sample Relinquished by: JD/CDC