


48126165

7471B

INSTRUMENT 101 CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.026	-0.026	236	0.000	236					-4
0.2	0.200	0.193	-0.007	1457	0.0 %	1457					-9
2.5	2.500	2.273	-0.227	13054	0.0 %	13054					-7
5.0	5.000	5.325	0.325	30069	0.0 %	30069					-4
7.5	7.500	7.830	0.330	44038	0.0 %	44038					-4
10.0	10.000	9.605	-0.395	53932	0.0 %	53932					

LB126165 INSTRUMENT ID CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Jun 2023 18:09:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	236	-	0.0000	7471B	PPB	23 Jun 2023 18:13:30	S	Std
0.2 - 1	50.2	1457	-	0.2000	7471B	PPB	23 Jun 2023 18:15:47	S	Std
2.5 - 1	52.5	13054	-	2.5000	7471B	PPB	23 Jun 2023 18:18:05	S	Std
5.0 - 1	55	30069	-	5.0000	7471B	PPB	23 Jun 2023 18:22:58	S	Std
7.5 - 1	57.5	44038	-	7.5000	7471B	PPB	23 Jun 2023 18:27:48	S	Std
10.0 - 1	510	53932	-	10.0000	7471B	PPB	23 Jun 2023 18:32:41	S	Std
ICV39 - 1	ICV39	22292	3.9300	-7471B	PPB	23 Jun 2023 18:38:04	U	SMPL	
ICB39 - 1	ICB39	61	-0.0572	-7471B	PPB	23 Jun 2023 18:40:20	U	SMPL	
CCV23 - 1	CCV23	29612	5.2428	-7471B	PPB	23 Jun 2023 18:42:37	U	SMPL	
CCB23 - 1	CCB23	64	-0.0566	-7471B	PPB	23 Jun 2023 18:44:52	U	SMPL	
CRA - 1	CRA	1429	0.1882	-7471B	PPB	23 Jun 2023 18:47:09	U	SMPL	
HighStd - 1	HighStd	54093	9.6335	-7471B	PPB	23 Jun 2023 18:49:24	U	SMPL	
ChkStd - 1	ChkStd	41969	7.4590	-7471B	PPB	23 Jun 2023 18:51:38	U	SMPL	
PB153763BL - 1	PBS	140	-0.0430	-7471B	PPB	23 Jun 2023 18:53:56	U	SMPL	
PB153763BS - 1	LCSS	21115	3.7189	-7471B	PPB	23 Jun 2023 18:56:14	U	SMPL	
O3375-01 - 1	SB-5(0-2)	30570	5.4146	-7471B	PPB	23 Jun 2023 18:58:30	U	SMPL	
O3375-02 - 1	SB-5(8-10)	521	0.0253	-7471B	PPB	23 Jun 2023 19:00:48	U	SMPL	
O3375-03 - 1	SB-2(0-2)	48040	8.5479	-7471B	PPB	23 Jun 2023 19:03:07	U	SMPL	
O3375-04 - 1	SB-2(8-10)	1919	0.2761	-7471B	PPB	23 Jun 2023 19:05:24	U	SMPL	
O3375-04DUP - 1	SB-2(8-10)DUP	2130	0.3139	-7471B	PPB	23 Jun 2023 19:10:17	U	SMPL	
CCV24 - 1	CCV24	28559	5.0539	-7471B	PPB	23 Jun 2023 19:12:32	U	SMPL	
CCB24 - 1	CCB24	129	-0.0450	-7471B	PPB	23 Jun 2023 19:14:47	U	SMPL	
O3375-05 - 1	O3375-04MS	24120	4.2578	-7471B	PPB	23 Jun 2023 19:17:03	U	SMPL	
O3375-06 - 1	O3375-04MSD	26347	4.6572	-7471B	PPB	23 Jun 2023 19:19:18	U	SMPL	
O3375-07 - 1	SB-3(0-2)	66106	11.7880	-7471B	PPB	23 Jun 2023 19:21:35	U	SMPL	
O3375-08 - 1	SB-3(8-10)	3354	0.5334	-7471B	PPB	23 Jun 2023 19:23:53	U	SMPL	
O3388-01 - 1	TR-05-062223	8513	1.4587	-7471B	PPB	23 Jun 2023 19:26:11	U	SMPL	
O3392-01 - 1	RY-1	34864	6.1847	-7471B	PPB	23 Jun 2023 19:28:28	U	SMPL	
O3392-03 - 1	SRY-1	3803	0.6140	-7471B	PPB	23 Jun 2023 19:30:44	U	SMPL	
O3392-05 - 1	SRY-3	5029	0.8338	-7471B	PPB	23 Jun 2023 19:33:03	U	SMPL	
O3392-07 - 1	SB-64	8239	1.4096	-7471B	PPB	23 Jun 2023 19:35:20	U	SMPL	
O3393-01 - 1	SF-9	14842	2.5938	-7471B	PPB	23 Jun 2023 19:37:36	U	SMPL	
CCV25 - 1	CCV25	28829	5.1024	-7471B	PPB	23 Jun 2023 19:39:52	U	SMPL	
CCB25 - 1	CCB25	152	-0.0409	-7471B	PPB	23 Jun 2023 19:42:08	U	SMPL	
O3393-03 - 1	SF-10	5934	0.9961	-7471B	PPB	23 Jun 2023 19:44:25	U	SMPL	
O3394-01 - 1	ORA-1651	19079	3.3537	-7471B	PPB	23 Jun 2023 19:46:41	U	SMPL	
O3398-01 - 1	WC-1	4463	0.7323	-7471B	PPB	23 Jun 2023 19:48:56	U	SMPL	
O3375-04LX5 - 1		339	-0.0073	-7471B	PPB	23 Jun 2023 19:51:14	U	SMPL	
O3375-04A - 1		23805	4.2013	-7471B	PPB	23 Jun 2023 19:53:30	U	SMPL	
O3375-07DLX2 - 1		33448	5.9308	-7471B	PPB	23 Jun 2023 19:55:46	U	SMPL	
CCV26 - 1	CCV26	28937	5.1217	-7471B	PPB	23 Jun 2023 19:58:03	U	SMPL	
CCB26 - 1	CCB26	87	-0.0525	-7471B	PPB	23 Jun 2023 20:00:22	U	SMPL	