


L3126170

SFAM01.1

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.9456e-004 slope

C= 9.1449e-003 $\rightarrow$ to intercept

Rho= 0.9997603

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	b/d
0.00	0.000	0.061	0.061	264	0.000	264					
0.05	0.050					302					
0.20	0.200	0.184	-0.016	901	0.0 %	901	0				
2.50	2.500	2.467	-0.033	12632	0.0 %	12632					
5.00	5.000	4.895	-0.105	25114	0.0 %	25114					
7.50	7.500	7.644	0.144	39240	0.0 %	39240					
10.0	10.000	9.949	-0.051	51091	0.0 %	51091					

b/d
= 8
- 1
- 2
- 2
- 1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	264	-		0 SFAM01.1	PPB	6/24/2023 11:01	Std
	0.2 S01	901	-		0.2 SFAM01.1	PPB	6/24/2023 11:03	Std
	2.5 S02	12632	-		2.5 SFAM01.1	PPB	6/24/2023 11:05	Std
	5 S03	25114	-		5 SFAM01.1	PPB	6/24/2023 11:08	Std
	7.5 S04	39240	-		7.5 SFAM01.1	PPB	6/24/2023 11:10	Std
	10 S05	51091	-		10 SFAM01.1	PPB	6/24/2023 11:15	Std
ICV040	ICV040	20142	3.9279 -		SFAM01.1	PPB	6/24/2023 11:21	SMPL
ICB040	ICB040	157	0.0397 -		SFAM01.1	PPB	6/24/2023 11:23	SMPL
CCV027	CCV027	24286	4.7342 -		SFAM01.1	PPB	6/24/2023 11:25	SMPL
CCB027	CCB027	89	0.0265 -		SFAM01.1	PPB	6/24/2023 11:27	SMPL
PB153774BL	PBW774	66	0.022 -		SFAM01.1	PPB	6/24/2023 11:30	SMPL
O3099-01	MEZEN4	158	0.0399 -		SFAM01.1	PPB	6/24/2023 11:32	SMPL
O3099-02	MEZEN4D	203	0.0486 -		SFAM01.1	PPB	6/24/2023 11:34	SMPL
O3099-03	MEZEN4S	5059	0.9934 -		SFAM01.1	PPB	6/24/2023 11:36	SMPL
O3099-04	MEZEN2	108	0.0302 -		SFAM01.1	PPB	6/24/2023 11:39	SMPL
O3099-05	MEZEN3	384	0.0839 -		SFAM01.1	PPB	6/24/2023 11:41	SMPL
O3099-06	MEZEP1	194	0.0469 -		SFAM01.1	PPB	6/24/2023 11:43	SMPL
O3099-07	MEZEP5	145	0.0374 -		SFAM01.1	PPB	6/24/2023 11:45	SMPL
O3099-08	MEZEN6	124	0.0333 -		SFAM01.1	PPB	6/24/2023 11:48	SMPL
O3099-09	MEZEP0	145	0.0374 -		SFAM01.1	PPB	6/24/2023 11:50	SMPL
O3099-10	MEZEN7	110	0.0305 -		SFAM01.1	PPB	6/24/2023 11:52	SMPL
O3099-11	MEZEN8	96	0.0278 -		SFAM01.1	PPB	6/24/2023 11:54	SMPL
O3099-12	MEZEQ6	111	0.0307 -		SFAM01.1	PPB	6/24/2023 11:57	SMPL
O3099-13	MEZERO	173	0.0428 -		SFAM01.1	PPB	6/24/2023 11:59	SMPL
O3099-14	MEZEP6	219	0.0518 -		SFAM01.1	PPB	6/24/2023 12:01	SMPL
O3099-15	MEZEQ5	95	0.0276 -		SFAM01.1	PPB	6/24/2023 12:03	SMPL
PB153775BL	PBW775	112	0.0309 -		SFAM01.1	PPB	6/24/2023 12:06	SMPL
O3100-01	MEZEP9	102	0.029 -		SFAM01.1	PPB	6/24/2023 12:08	SMPL
O3100-02	MEZEP9D	186	0.0453 -		SFAM01.1	PPB	6/24/2023 12:10	SMPL
O3100-03	MEZEP9S	4274	0.8407 -		SFAM01.1	PPB	6/24/2023 12:13	SMPL
CCV028	CCV028	24885	4.8507 -		SFAM01.1	PPB	6/24/2023 12:20	SMPL
CCB028	CCB028	59	0.0206 -		SFAM01.1	PPB	6/24/2023 12:22	SMPL
O3100-04	MEZEQ1	92	0.027 -		SFAM01.1	PPB	6/24/2023 12:25	SMPL
O3100-05	MEZEQ7	101	0.0288 -		SFAM01.1	PPB	6/24/2023 12:27	SMPL
O3100-06	MEZEP8	89	0.0265 -		SFAM01.1	PPB	6/24/2023 12:29	SMPL
O3100-07	MEZEP7	86	0.0259 -		SFAM01.1	PPB	6/24/2023 12:32	SMPL
O3100-08	MEZEQ0	181	0.0444 -		SFAM01.1	PPB	6/24/2023 12:34	SMPL
PB153776BL	PBW776	125	0.0335 -		SFAM01.1	PPB	6/24/2023 12:36	SMPL
O3112-07	X3G75	12223	2.3872 -		SFAM01.1	PPB	6/24/2023 12:41	SMPL
O3112-08	X3G75D	12259	2.3942 -		SFAM01.1	PPB	6/24/2023 12:43	SMPL
O3112-09	X3G75S	17276	3.3703 -		SFAM01.1	PPB	6/24/2023 12:47	SMPL
CCV029	CCV029	23353	4.5527 -		SFAM01.1	PPB	6/24/2023 12:52	SMPL
CCB029	CCB029	61	0.021 -		SFAM01.1	PPB	6/24/2023 12:54	SMPL