



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/26/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:05
In Date: 06/24/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 06/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB126172

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3349-01	YBTZ5	1	1.14	8.61	9.75	9.08	92.2	
O3349-02	YBTZ6	2	1.12	8.54	9.66	8.01	80.7	
O3349-03	YBTZ7	3	1.15	8.44	9.59	7.45	74.6	
O3349-04	YBTZ9	4	1.13	8.62	9.75	7.68	76.0	
O3349-05	YBW00	5	1.15	8.73	9.88	8.31	82.0	
O3349-06	YBW01	6	1.14	8.67	9.81	8.25	82.0	
O3349-07	YBW01MS	7	1.14	8.67	9.81	8.25	82.0	
O3349-08	YBW01MSD	8	1.14	8.67	9.81	8.25	82.0	
O3349-09	YBW02	9	1.11	8.66	9.77	7.92	78.6	
O3349-10	YBW03	10	1.14	8.70	9.84	7.72	75.6	
O3349-11	YBW04	11	1.13	8.66	9.79	9.44	96.0	
O3349-12	YBW05	12	1.13	8.45	9.58	9.33	97.0	
O3349-13	YBW06	13	1.14	8.49	9.63	8.25	83.7	
O3349-14	YBW07	14	1.12	8.50	9.62	8.51	86.9	
O3349-15	YBW08	15	1.12	8.63	9.75	8.54	86.0	
O3349-16	YBW09	16	1.12	8.40	9.52	8.81	91.5	
O3349-17	YBW11	17	1.14	8.36	9.5	8.29	85.5	
O3349-18	YBW12	18	1.11	8.57	9.68	8.43	85.4	
O3349-19	YBW22	19	1.19	8.60	9.79	8.56	85.7	
O3349-20	YBW23	20	1.11	8.57	9.68	9.38	96.5	
O3349-21	YBW24	21	1.16	8.54	9.7	8.01	80.2	
O3349-22	YBW25	22	1.11	8.87	9.98	8.07	78.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$