

LB126127


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

65322



A= 0.0000e+000

B= 1.5627e-004

C= 4.3583e-002

Rho= 0.9983990

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	40D
0.0	0.000	0.094	0.094	324	0.000	324					14
0.2	0.200	0.228	0.028	1182	0.0 %	1182					7
2.5	2.500	2.670	0.170	16807	0.0 %	16807					-7
5.0	5.000	4.651	-0.349	29482	0.0 %	29482					-3
7.5	7.500	7.306	-0.194	46474	0.0 %	46474					3
10.0	10.000	10.251	0.251	65322	0.0 %	65322					

LB126187 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jun 2023 15:03:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	324	-	0.0000	7471B	PPB	26 Jun 2023 15:06:34	S	Std
0.2 - 1	50.2	1182	-	0.2000	7471B	PPB	26 Jun 2023 15:08:51	S	Std
2.5 - 1	52.5	16807	-	2.5000	7471B	PPB	26 Jun 2023 15:11:08	S	Std
5.0 - 1	55	29482	-	5.0000	7471B	PPB	26 Jun 2023 15:13:25	S	Std
7.5 - 1	57.5	46474	-	7.5000	7471B	PPB	26 Jun 2023 15:15:42	S	Std
10.0 - 1	51V	65322	-	10.0000	7471B	PPB	26 Jun 2023 15:31:02	S	Std
ICV42 - 1	ICV42	24063	3.8038	-	7471B	PPB	26 Jun 2023 15:33:43	U	SMPL
ICB42 - 1	ICB42	-59	0.0344	-	7471B	PPB	26 Jun 2023 15:35:59	U	SMPL
CCV34 - 1	CCV34	29833	4.7054	-	7471B	PPB	26 Jun 2023 15:38:16	U	SMPL
CCB34 - 1	CCB34	-70	0.0326	-	7471B	PPB	26 Jun 2023 15:40:31	U	SMPL
CRA - 1	CRA	1292	0.2455	-	7471B	PPB	26 Jun 2023 15:42:47	U	SMPL
HighStd - 1	HighStd	60289	9.4646	-	7471B	PPB	26 Jun 2023 15:45:02	U	SMPL
ChkStd - 1	ChkStd	41397	6.5125	-	7471B	PPB	26 Jun 2023 15:47:18	U	SMPL
PB153805BL - 1	PBS	-147	0.0206	-	7471B	PPB	26 Jun 2023 15:49:36	U	SMPL
PB153805BS - 1	LCSS	22352	3.5364	-	7471B	PPB	26 Jun 2023 15:51:54	U	SMPL
O3289-12 - 1	SB-05-0-2.0	9564	1.5381	-	7471B	PPB	26 Jun 2023 15:54:09	U	SMPL
O3402-01 - 1	TP1-0623	2090	0.3702	-	7471B	PPB	26 Jun 2023 15:56:27	U	SMPL
O3402-01DUP - 1	TP1-0623DUP	2214	0.3896	-	7471B	PPB	26 Jun 2023 15:58:44	U	SMPL
O3402-01MS - 1	TP1-0623MS	29775	4.6964	-	7471B	PPB	26 Jun 2023 16:01:01	U	SMPL
O3402-01MSD - 1	TP1-0623MSD	27946	4.4106	-	7471B	PPB	26 Jun 2023 16:03:17	U	SMPL
CCV35 - 1	CCV35	29908	4.7172	-	7471B	PPB	26 Jun 2023 16:05:33	U	SMPL
CCB35 - 1	CCB35	-33	0.0384	-	7471B	PPB	26 Jun 2023 16:10:26	U	SMPL
O3402-03 - 1	TP2-0623	1179	0.2278	-	7471B	PPB	26 Jun 2023 16:12:41	U	SMPL
O3402-05 - 1	TP3-0623	1404	0.2630	-	7471B	PPB	26 Jun 2023 16:14:56	U	SMPL
O3402-07 - 1	TP4-0623	1364	0.2567	-	7471B	PPB	26 Jun 2023 16:17:10	U	SMPL
O3403-01 - 1	ORA-1652	17108	2.7170	-	7471B	PPB	26 Jun 2023 16:19:26	U	SMPL
O3409-01 - 1	OE-1-062323	60373	9.4778	-	7471B	PPB	26 Jun 2023 16:21:42	U	SMPL
O3402-01LX5 - 1		219	0.0778	-	7471B	PPB	26 Jun 2023 16:23:59	U	SMPL
O3402-01A - 1		28575	4.5089	-	7471B	PPB	26 Jun 2023 16:26:18	U	SMPL
CCV36 - 1	CCV36	29780	4.6972	-	7471B	PPB	26 Jun 2023 16:28:35	U	SMPL
CCB36 - 1	CCB36	-142	0.0214	-	7471B	PPB	26 Jun 2023 16:30:54	U	SMPL