



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126190

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/26/2023	13:50
2	CAL2	1	Water	NA	NA	20.8	7.00	06/26/2023	13:51
3	CAL3	1	Water	NA	NA	20.8	10.02	06/26/2023	13:52
4	ICV	1	Water	NA	NA	20.7	6.99	06/26/2023	14:00
5	CCV1	1	Water	NA	NA	20.7	2.01	06/26/2023	14:02
6	O3417-01	1	Water	NA	NA	23.7	7.50	06/26/2023	14:10
7	O3417-02	1	Water	NA	NA	23.5	6.58	06/26/2023	14:15
8	O3417-03	1	Water	NA	NA	23.7	6.71	06/26/2023	14:20
9	O3417-03DUP	1	Water	NA	NA	23.8	6.72	06/26/2023	14:21
10	CCV2	1	Water	NA	NA	20.7	12.02	06/26/2023	14:25

VB 126190

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 03417 WorkList ID : 171345 Department : Wet-Chemistry Date : 06-26-2023 13:40:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3417-01	A 2700	Water	pH	Cool 4 deg C	PSEG03	I11	06/26/2023	9040C
O3417-02	C 1621	Water	pH	Cool 4 deg C	PSEG03	I11	06/26/2023	9040C
O3417-03	C 1622	Water	pH	Cool 4 deg C	PSEG03	I11	06/26/2023	9040C

Date/Time 06/26/23 13:45
Raw Sample Received by: JD LWCJ
Raw Sample Relinquished by: AL SM

Date/Time 06/26/23 16:00
Raw Sample Received by: AL SM
Raw Sample Relinquished by: JD LWCJ