



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/28/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 06/27/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 06/28/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126201

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3406-01	MJVA08	1	1.15	8.77	9.92	7.6	73.5	
O3406-02	MJVA09	2	1.15	8.82	9.97	7.27	69.4	
O3406-03	MJVA82	3	1.14	8.81	9.95	7.17	68.4	
O3406-04	MJVA06	4	1.15	8.84	9.99	5.89	53.6	
O3406-05	MJVA48	5	1.19	8.37	9.56	3.63	29.2	
O3406-06	MJVA50	6	1.18	8.64	9.82	5.02	44.4	
O3406-07	MJVA52	7	1.11	8.86	9.97	4.71	40.6	
O3406-08	MJVA53	8	1.12	8.64	9.76	4.75	42.0	
O3406-09	MJVA63	9	1.13	8.53	9.66	3.67	29.8	
O3406-10	MJVA64	10	1.15	8.54	9.69	5.09	46.1	
O3406-11	MJVA73	11	1.13	8.76	9.89	4.91	43.2	
O3406-12	MJVA73D	12	1.13	8.76	9.89	4.91	43.2	
O3406-13	MJVA73S	13	1.13	8.76	9.89	4.91	43.2	
O3406-14	MJVA74	14	1.17	8.69	9.86	5.74	52.6	
O3406-15	MJVA86	15	1.12	8.54	9.66	5.12	46.8	
O3406-16	MJVA54	16	1.19	8.80	9.99	6.4	59.2	
O3406-17	MJVA67	17	1.14	8.84	9.98	6.62	62.0	
O3406-18	MJVA81	18	1.16	8.70	9.86	8.00	78.6	
O3406-19	MJVA96	19	1.14	8.85	9.99	7.22	68.7	
O3406-20	MJVA97	20	1.18	8.78	9.96	7.03	66.6	
O3406-21	MJVAA6	21	1.16	8.49	9.65	6.62	64.3	
O3406-22	MJVAA7	22	1.13	8.77	9.9	8.37	82.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$