

LB126202


SFAM01.1

INSTRUMENT ID: CV3

Linear

 μ Abs.:

5855E



A= 0.0000e+000

B= 1.7078e-004 *slope*C= -7.9455e-002 *intercept*

Rho= 0.9998519

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	400
0.00	0.000	-0.002	-0.002	452	0.000	452					
0.05	0.050					302	0				
0.20	0.200	0.179	-0.021	1512	0.0 %	1512					
2.50	2.500	2.445	-0.055	14780	0.0 %	14780					
5.00	5.000	5.110	0.110	30386	0.0 %	30386					
7.50	7.500	7.548	0.048	44664	0.0 %	44664					
10.0	10.000	9.921	-0.079	58556	0.0 %	58556					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	452	-		0 SFAM01.1	PPB	6/27/2023 13:43	Std
	0.2 S01	1512	-		0.2 SFAM01.1	PPB	6/27/2023 13:46	Std
	2.5 S02	14780	-		2.5 SFAM01.1	PPB	6/27/2023 13:48	Std
	5 S03	30386	-		5 SFAM01.1	PPB	6/27/2023 13:50	Std
	7.5 S04	44664	-		7.5 SFAM01.1	PPB	6/27/2023 13:52	Std
	10 S05	58556	-		10 SFAM01.1	PPB	6/27/2023 13:55	Std
ICV043	ICV043	24396	4.0869 -		SFAM01.1	PPB	6/27/2023 14:00	SMPL
ICB043	ICB043	245	-0.0376 -		SFAM01.1	PPB	6/27/2023 14:02	SMPL
CCV037	CCV037	30176	5.074 -		SFAM01.1	PPB	6/27/2023 14:05	SMPL
CCB037	CCB037	255	-0.0359 -		SFAM01.1	PPB	6/27/2023 14:07	SMPL
PB153829BL	PBW829	111	-0.0605 -		SFAM01.1	PPB	6/27/2023 14:09	SMPL
O3285-01	MDCHC0	46100	7.7935 -		SFAM01.1	PPB	6/27/2023 14:11	SMPL
O3285-03	MDCGZ1	-78	-0.0928 -		SFAM01.1	PPB	6/27/2023 14:14	SMPL
O3285-04	MDCGZ2	108	-0.061 -		SFAM01.1	PPB	6/27/2023 14:16	SMPL
O3285-05	MDCGZ3	201	-0.0451 -		SFAM01.1	PPB	6/27/2023 14:18	SMPL
O3285-06	MDCGZ4	264	-0.0344 -		SFAM01.1	PPB	6/27/2023 14:21	SMPL
O3285-07	MDCGZ5	10179	1.6589 -		SFAM01.1	PPB	6/27/2023 14:23	SMPL
O3285-08	MDCGZ6	4	-0.0788 -		SFAM01.1	PPB	6/27/2023 14:25	SMPL
O3285-09	MDCGZ6D	190	-0.047 -		SFAM01.1	PPB	6/27/2023 14:27	SMPL
O3285-10	MDCGZ6S	6016	0.948 -		SFAM01.1	PPB	6/27/2023 14:30	SMPL
O3285-11	MDCGZ7	80	-0.0658 -		SFAM01.1	PPB	6/27/2023 14:32	SMPL
O3285-12	MDCGZ9	425	-0.0069 -		SFAM01.1	PPB	6/27/2023 14:34	SMPL
PB153830BL	PBW830	401	-0.011 -		SFAM01.1	PPB	6/27/2023 14:36	SMPL
O3323-01	MDCHA0	497	0.0054 -		SFAM01.1	PPB	6/27/2023 14:39	SMPL
O3323-02	MDCHA1	711	0.042 -		SFAM01.1	PPB	6/27/2023 14:41	SMPL
O3323-04	MDCHA3	615	0.0256 -		SFAM01.1	PPB	6/27/2023 14:43	SMPL
O3323-05	MDCHA4	271	-0.0332 -		SFAM01.1	PPB	6/27/2023 14:45	SMPL
O3323-06	MDCHA5	300	-0.0282 -		SFAM01.1	PPB	6/27/2023 14:48	SMPL
O3323-07	MDCHA6	260	-0.0351 -		SFAM01.1	PPB	6/27/2023 14:50	SMPL
O3323-08	MDCHA6D	320	-0.0248 -		SFAM01.1	PPB	6/27/2023 14:52	SMPL
CCV038	CCV038	29458	4.9514 -		SFAM01.1	PPB	6/27/2023 14:55	SMPL
CCB038	CCB038	334	-0.0224 -		SFAM01.1	PPB	6/27/2023 14:57	SMPL
O3323-09	MDCHA6S	6074	0.9579 -		SFAM01.1	PPB	6/27/2023 14:59	SMPL
O3323-10	MDCHA7	395	-0.012 -		SFAM01.1	PPB	6/27/2023 15:01	SMPL
CCV039	CCV039	29306	4.9254 -		SFAM01.1	PPB	6/27/2023 15:11	SMPL
CCB039	CCB039	221	-0.0417 -		SFAM01.1	PPB	6/27/2023 15:13	SMPL