



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126205

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/27/2023	15:45
2	CAL2	1	Water	NA	NA	20.7	7.00	06/27/2023	15:46
3	CAL3	1	Water	NA	NA	20.7	10.02	06/27/2023	15:47
4	ICV	1	Water	NA	NA	20.6	6.99	06/27/2023	15:50
5	CCV1	1	Water	NA	NA	20.6	2.01	06/27/2023	16:00
6	O3430-01	1	Water	NA	NA	25.9	6.43	06/27/2023	16:02
7	O3430-03	1	Water	NA	NA	26.0	6.32	06/27/2023	16:10
8	O3430-03DUP	1	Water	NA	NA	26.1	6.33	06/27/2023	16:11
9	CCV2	1	Water	NA	NA	20.8	12.02	06/27/2023	16:15