

LB126206


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

70827



A= 0.0000e+000

B= 1.4098e-004

C= -3.2155e-002

Rho= 0.9998006

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.057	0.057	629	0.000	629					
0.2	0.200	0.216	0.016	1759	0.0 %	1759					
2.5	2.500	2.363	-0.137	16991	0.0 %	16991					
5.0	5.000	5.024	0.024	35863	0.0 %	35863					
7.5	7.500	7.588	0.088	54048	0.0 %	54048					
10.0	10.000	9.953	-0.047	70827	0.0 %	70827					

9.11
18
75
10
0

LB126206 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Jun 2023 15:55:28

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	629	-	0.0000	7471B	PPB	27 Jun 2023 16:06:01	S	Std
0.2 - 1	50.2	1759	-	0.2000	7471B	PPB	27 Jun 2023 16:08:18	S	Std
2.5 - 1	52.5	16991	-	2.5000	7471B	PPB	27 Jun 2023 16:10:35	S	Std
5.0 - 1	55	35863	-	5.0000	7471B	PPB	27 Jun 2023 16:12:52	S	Std
7.5 - 1	57.5	54048	-	7.5000	7471B	PPB	27 Jun 2023 16:15:08	S	Std
10.0 - 1	511	70827	-	10.0000	7471B	PPB	27 Jun 2023 16:17:26	S	Std
ICV44 - 1	ICV44	30085	4.2092	-	7471B	PPB	27 Jun 2023 16:20:14	U	SMPL
ICB44 - 1	ICB44	494	0.0375	-	7471B	PPB	27 Jun 2023 16:22:29	U	SMPL
CCV40 - 1	CCV40	35822	5.0180	-	7471B	PPB	27 Jun 2023 16:24:46	U	SMPL
CCB40 - 1	CCB40	506	0.0392	-	7471B	PPB	27 Jun 2023 16:27:01	U	SMPL
CRA - 1	CRA	1533	0.1840	-	7471B	PPB	27 Jun 2023 16:29:18	U	SMPL
HighStd - 1	HighStd	70253	9.8721	-	7471B	PPB	27 Jun 2023 16:31:33	U	SMPL
ChkStd - 1	ChkStd	50124	7.0343	-	7471B	PPB	27 Jun 2023 16:33:48	U	SMPL
PB153832BL - 1	PBS	165	-0.0089	-	7471B	PPB	27 Jun 2023 16:36:06	U	SMPL
PB153832BS - 1	LCSS	25312	3.5363	-	7471B	PPB	27 Jun 2023 16:38:25	U	SMPL
O3413-01 - 1	OR-02-062623	5286	0.7131	-	7471B	PPB	27 Jun 2023 16:40:41	U	SMPL
O3413-01DUP - 1	OR-02-062623DUP	4674	0.6268	-	7471B	PPB	27 Jun 2023 16:42:58	U	SMPL
O3413-01MS - 1	OR-02-062623MS	31383	4.3922	-	7471B	PPB	27 Jun 2023 16:45:15	U	SMPL
O3413-01MSD - 1	OR-02-062623MSD	33575	4.7013	-	7471B	PPB	27 Jun 2023 16:47:32	U	SMPL
O3415-08 - 1	ESII-TP1	4818	0.6471	-	7471B	PPB	27 Jun 2023 16:49:49	U	SMPL
CCV41 - 1	CCV41	35650	4.9938	-	7471B	PPB	27 Jun 2023 16:52:06	U	SMPL
CCB41 - 1	CCB41	561	0.0469	-	7471B	PPB	27 Jun 2023 16:54:21	U	SMPL
O3428-01 - 1	PNNL-APM-0621-1	3397	0.4468	-	7471B	PPB	27 Jun 2023 16:56:37	U	SMPL
O3428-03 - 1	PNNL-APM-0621-2	4255	0.5677	-	7471B	PPB	27 Jun 2023 16:58:53	U	SMPL
O3413-01LX5 - 1		867	0.0901	-	7471B	PPB	27 Jun 2023 17:01:08	U	SMPL
O3413-01A - 1		32488	4.5480	-	7471B	PPB	27 Jun 2023 17:03:24	U	SMPL
CCV42 - 1	CCV42	35905	5.0297	-	7471B	PPB	27 Jun 2023 17:05:39	U	SMPL
CCB42 - 1	CCB42	559	0.0467	-	7471B	PPB	27 Jun 2023 17:07:57	U	SMPL