



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/28/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:55
In Date: 06/27/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 06/28/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126207

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3428-01	PNNL-APM-0621-1	1	1.14	8.76	9.9	9.12	91.1	
O3428-03	PNNL-APM-0621-2	2	1.19	8.57	9.76	9.12	92.5	
O3429-01	SB-COMP-0627	3	1.19	8.41	9.6	8.78	90.2	
O3429-03	SB-COMP-0627-EPH-2	4	1.11	8.73	9.84	9.28	93.6	
O3431-01	40249	5	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-02	40250	6	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-03	40251	7	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-04	40252	8	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-05	40253	9	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-06	40254	10	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-07	40255	11	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-08	40256	12	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-09	40257	13	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-10	40258	14	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-11	40259	15	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-12	40260	16	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-13	40261	17	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3431-14	40262	18	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3434-01	RT2665	19	1.15	8.58	9.73	9.06	92.2	
O3434-02	RT2665	20	1.11	8.61	9.72	9.1	92.8	
O3434-03	350	21	1.12	8.56	9.68	9.14	93.7	
O3434-04	350	22	1.12	8.71	9.83	9.03	90.8	
O3435-01	40044	23	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3436-01	TRE-1754	24	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3441-01	KL-FM628-F06064	25	1.19	8.79	9.98	4.02	32.2	SLUDGE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$