



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126222

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	06/28/2023	08:10
2	O3430-01		1	12	100.00	>212.0	>212.0	06/28/2023	08:40
3	O3430-01DUP		1	14	100.00	>212.0	>212.0	06/28/2023	09:10
4	O3430-03		1	13	100.00	>212.0	>212.0	06/28/2023	09:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

25126222

WorkList Name : fp 6-27

WorkList ID : 171390

Department : Wet-Chemistry

Date : 06-27-2023 15:43:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3430-01	1003	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/27/2023	1010B
O3430-03	1004	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/27/2023	1010B

Date/Time 06/28/2023 08:00
 Raw Sample Received by: Rm (wv)
 Raw Sample Relinquished by: Rm (wv)

Date/Time 06/28/2023 10:20
 Raw Sample Received by: Rm (wv)
 Raw Sample Relinquished by: Rm (wv)