



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126230

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/28/2023	13:10
2	CAL2	1	Water	NA	NA	20.8	7.00	06/28/2023	13:11
3	CAL3	1	Water	NA	NA	20.8	10.02	06/28/2023	13:15
4	ICV	1	Water	NA	NA	20.8	6.99	06/28/2023	13:20
5	CCV1	1	Water	NA	NA	20.7	2.01	06/28/2023	13:25
6	O3456-01	1	Water	NA	NA	23.9	4.23	06/28/2023	13:30
7	O3456-01DUP	1	Water	NA	NA	23.8	4.24	06/28/2023	13:31
8	O3456-02	1	Water	NA	NA	23.7	4.26	06/28/2023	13:35
9	CCV2	1	Water	NA	NA	20.6	12.02	06/28/2023	13:40