



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB126234

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/28/2023	17:15
2	CAL2	1	Water	NA	NA	20.7	7.00	06/28/2023	17:16
3	CAL3	1	Water	NA	NA	20.7	10.03	06/28/2023	17:19
4	ICV	1	Water	NA	NA	20.6	6.99	06/28/2023	17:20
5	CCV1	1	Water	NA	NA	20.8	2.01	06/28/2023	17:50
6	O3464-03	1	Water	NA	NA	25.5	6.27	06/28/2023	18:00
7	O3464-03DUP	1	Water	NA	NA	25.6	6.28	06/28/2023	18:01
8	O3464-04	1	Water	NA	NA	25.6	6.20	06/28/2023	18:05
9	O3465-03	1	Water	NA	NA	26.5	7.54	06/28/2023	18:10
10	O3465-04	1	Water	NA	NA	26.3	6.07	06/28/2023	18:15
11	O3471-05	1	Water	NA	NA	25.9	7.72	06/28/2023	18:20
12	CCV2	1	Water	NA	NA	20.6	12.02	06/28/2023	18:22

YB 126234

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity 03471 W

WorkList ID : 171448

Department : Wet-Chemistry

Date : 06-28-2023 17:01:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3464-03	366	Water	Corrosivity	Cool 4 deg C	PSEG03	I11	06/28/2023	9040C
O3464-04	367	Water	Corrosivity	Cool 4 deg C	PSEG03	I11	06/28/2023	9040C
O3465-03	3110	Water	Corrosivity	Cool 4 deg C	PSEG03	I11	06/28/2023	9040C
O3465-04	3041-3043	Water	Corrosivity	Cool 4 deg C	PSEG03	I11	06/28/2023	9040C
O3471-05	204	Water	Corrosivity	Cool 4 deg C	PSEG03	I11	06/28/2023	9040C

Date/Time 06/28/23 17:10
 Raw Sample Received by: JB (60C)
 Raw Sample Relinquished by: RS CERT-100

Date/Time 06/28/23 19:00
 Raw Sample Received by: RS CERT-100
 Raw Sample Relinquished by: JB (60C)