



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/3/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 16:10

In Date: 06/30/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:15

Out Date: 07/01/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: % SOLIDS-OVEN

QC:LB126239

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3420-01	A4636	1	1.15	8.82	9.97	8.16	79.5	
O3420-02	A4637	2	1.19	8.57	9.76	8.53	85.6	
O3420-03	A4639	3	1.17	8.80	9.97	7.46	71.5	
O3420-04	A4640	4	1.19	8.51	9.7	8.97	91.4	
O3420-05	A4641	5	1.11	8.47	9.58	8.64	88.9	
O3420-06	A4645	6	1.13	8.85	9.98	4.43	37.3	
O3420-07	A4646	7	1.16	8.50	9.66	4.36	37.6	
O3420-08	A4647	8	1.19	8.60	9.79	4.64	40.1	
O3420-09	A4648	9	1.00	1.00	2.00	2.00	100.0	F.B.
O3420-10	A4652	10	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3420-11	A46E5	11	1.15	8.80	9.95	2.86	19.4	
O3420-12	A46E6	12	1.19	8.78	9.97	7.5	71.9	
O3420-13	A46E8	13	1.19	8.61	9.8	5.45	49.5	
O3420-14	A46E9	14	1.13	8.66	9.79	6.11	57.5	
O3420-15	A46F0	15	1.15	8.80	9.95	7.38	70.8	
O3420-16	A46F1	16	1.13	8.62	9.75	8.55	86.1	
O3420-17	A46G8	17	1.14	8.79	9.93	9.25	92.3	
O3420-18	A46G9	18	1.18	8.81	9.99	9.1	89.9	
O3420-19	A46H0	19	1.13	8.37	9.5	8.64	89.7	
O3420-20	VHBLK001	20	1.00	1.00	2.00	2.00	100.0	VHBLK
O3420-21	A46K5	21	1.16	8.82	9.98	7.59	72.9	
O3420-22	A46K5MS	22	1.16	8.82	9.98	7.59	72.9	
O3420-23	A46K5MSD	23	1.16	8.82	9.98	7.59	72.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$