

LB126245


SFAM01.1

INSTRUMENT ID: CVJ

Linear

μ Abs.:

57272



A= 0.0000e+000

B= 1.7376e-004 slope

C= 3.0312e-002 y-intercept

Rho= 0.9999100

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/6D
0.00	0.000	0.042	0.042	66	0.000	66					=
0.05	0.050					302	0				+6
0.20	0.200	0.212	0.012	1047	0.0 %	1047					-2
2.50	2.500	2.440	-0.060	13868	0.0 %	13868					-1
5.00	5.000	4.947	-0.053	28294	0.0 %	28294					-1
7.50	7.500	7.578	0.078	43436	0.0 %	43436					0
10.0	10.000	9.982	-0.018	57272	0.0 %	57272					

LB126245

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	66	-		0 SFAM01.1	PPB	6/29/2023 13:24	Std
	0.2 S01	1047	-		0.2 SFAM01.1	PPB	6/29/2023 13:27	Std
	2.5 S02	13868	-		2.5 SFAM01.1	PPB	6/29/2023 13:29	Std
	5 S03	28294	-		5 SFAM01.1	PPB	6/29/2023 13:31	Std
	7.5 S04	43436	-		7.5 SFAM01.1	PPB	6/29/2023 13:39	Std
	10 S05	57272	-		10 SFAM01.1	PPB	6/29/2023 13:41	Std
ICV047	ICV047	23956	4.1928 -		SFAM01.1	PPB	6/29/2023 13:44	SMPL
ICB047	ICB047	-75	0.0173 -		SFAM01.1	PPB	6/29/2023 13:47	SMPL
CCV052	CCV052	28576	4.9956 -		SFAM01.1	PPB	6/29/2023 13:49	SMPL
CCB052	CCB052	-89	0.0148 -		SFAM01.1	PPB	6/29/2023 13:51	SMPL
PB153893BL	PBW893	92	0.0463 -		SFAM01.1	PPB	6/29/2023 13:53	SMPL
O3407-02	MJVA83	86	0.0453 -		SFAM01.1	PPB	6/29/2023 13:56	SMPL
O3407-03	MJVA88	123	0.0517 -		SFAM01.1	PPB	6/29/2023 13:58	SMPL
O3407-04	MJVA95	121	0.0513 -		SFAM01.1	PPB	6/29/2023 14:00	SMPL
O3407-05	MJVA98	97	0.0472 -		SFAM01.1	PPB	6/29/2023 14:03	SMPL
O3407-06	MJVA99	124	0.0519 -		SFAM01.1	PPB	6/29/2023 14:05	SMPL
O3407-07	MJVAA0	441	0.1069 -		SFAM01.1	PPB	6/29/2023 14:12	SMPL
O3407-08	MJVAA0D	355	0.092 -		SFAM01.1	PPB	6/29/2023 14:14	SMPL
O3407-09	MJVAA0S	6662	1.1879 -		SFAM01.1	PPB	6/29/2023 14:16	SMPL
O3407-10	MJVAA2	233	0.0708 -		SFAM01.1	PPB	6/29/2023 14:18	SMPL
O3407-11	MJVAA3	266	0.0765 -		SFAM01.1	PPB	6/29/2023 14:21	SMPL
O3407-12	MJVAA4	87	0.0454 -		SFAM01.1	PPB	6/29/2023 14:23	SMPL
O3407-13	MJVAA5	523	0.1212 -		SFAM01.1	PPB	6/29/2023 14:25	SMPL
O3407-14	MJVAA8	28	0.0352 -		SFAM01.1	PPB	6/29/2023 14:27	SMPL
O3407-15	MJVAA9	25	0.0347 -		SFAM01.1	PPB	6/29/2023 14:30	SMPL
CCV053	CCV053	28196	4.9296 -		SFAM01.1	PPB	6/29/2023 14:32	SMPL
CCB053	CCB053	-20	0.0268 -		SFAM01.1	PPB	6/29/2023 14:34	SMPL