

LB126250


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

7090E



A= 0.0000e+000

B= 1.4014e-004

C= -6.0630e-002

Rho= 0.9997375

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/p</i>
0.0	0.000	0.018	0.018	559	0.000	559					-
0.2	0.200	0.162	-0.038	1587	0.0 %	1587					-19
2.5	2.500	2.463	-0.037	18004	0.0 %	18004					-1
5.0	5.000	5.027	0.027	36305	0.0 %	36305					1
7.5	7.500	7.654	0.154	55049	0.0 %	55049					2
10.0	10.000	9.877	-0.123	70908	0.0 %	70908					-1

LB126250 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Jun 2023 15:36:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	559	-	0.0000	7471B	PPB	29 Jun 2023 15:38:40	S	Std
0.2 - 1	51.2	1587	-	0.2000	7471B	PPB	29 Jun 2023 15:40:56	S	Std
2.5 - 1	52.8	18004	-	2.5000	7471B	PPB	29 Jun 2023 15:43:14	S	Std
5.0 - 1	55	36305	-	5.0000	7471B	PPB	29 Jun 2023 15:45:30	S	Std
7.5 - 1	57.5	55049	-	7.5000	7471B	PPB	29 Jun 2023 15:47:47	S	Std
10.0 - 1	58	70908	-	10.0000	7471B	PPB	29 Jun 2023 15:50:04	S	Std
ICV48 - 1	ICV48	28677	3.9583	-	7471B	PPB	29 Jun 2023 15:52:55	U	SMPL
ICB48 - 1	ICB48	520	0.0122	-	7471B	PPB	29 Jun 2023 15:55:09	U	SMPL
CCV54 - 1	CCV54	36572	5.0647	-	7471B	PPB	29 Jun 2023 15:57:26	U	SMPL
CCB54 - 1	CCB54	525	0.0129	-	7471B	PPB	29 Jun 2023 15:59:40	U	SMPL
CRA - 1	CRA	1826	0.1953	-	7471B	PPB	29 Jun 2023 16:04:28	U	SMPL
HighStd - 1	HighStd	70863	9.8703	-	7471B	PPB	29 Jun 2023 16:06:43	U	SMPL
ChkStd - 1	ChkStd	50155	6.9682	-	7471B	PPB	29 Jun 2023 16:08:58	U	SMPL
PB153900BL - 1	PBS	-74	-0.0710	-	7471B	PPB	29 Jun 2023 16:11:15	U	SMPL
PB153900BS - 1	LCSS	28839	3.9810	-	7471B	PPB	29 Jun 2023 16:16:06	U	SMPL
O3429-01 - 1	SB-COMP-0627	6313	0.8241	-	7471B	PPB	29 Jun 2023 16:18:22	U	SMPL
O3434-02 - 1	RT2665	29519	4.0763	-	7471B	PPB	29 Jun 2023 16:25:34	U	SMPL
O3434-02DUP - 1	RT2665DUP	29797	4.1152	-	7471B	PPB	29 Jun 2023 16:27:50	U	SMPL
O3434-02MS - 1	RT2665MS	58034	8.0724	-	7471B	PPB	29 Jun 2023 16:30:08	U	SMPL
O3434-02MSD - 1	RT2665MSD	52802	7.3392	-	7471B	PPB	29 Jun 2023 16:35:03	U	SMPL
CCV55 - 1	CCV55	36044	4.9907	-	7471B	PPB	29 Jun 2023 16:37:19	U	SMPL
CCB55 - 1	CCB55	519	0.0121	-	7471B	PPB	29 Jun 2023 16:39:36	U	SMPL
O3434-04 - 1	350	4409	0.5573	-	7471B	PPB	29 Jun 2023 16:41:54	U	SMPL
O3441-01 - 1	KL-FM628-F06064	64370	8.9604	-	7471B	PPB	29 Jun 2023 16:44:09	U	SMPL
O3462-01 - 1	SU-02-62823	508	0.0106	-	7471B	PPB	29 Jun 2023 16:46:24	U	SMPL
O3471-03 - 1	VNJ-232	3492	0.4287	-	7471B	PPB	29 Jun 2023 16:48:43	U	SMPL
O3434-02LX5 - 1		5953	0.7736	-	7471B	PPB	29 Jun 2023 16:50:59	U	SMPL
O3434-02A - 1		57428	7.9875	-	7471B	PPB	29 Jun 2023 16:53:15	U	SMPL
CCV56 - 1	CCV56	36731	5.0870	-	7471B	PPB	29 Jun 2023 16:55:32	U	SMPL
CCB56 - 1	CCB56	532	0.0139	-	7471B	PPB	29 Jun 2023 16:57:55	U	SMPL