



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126252

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	06/29/2023	11:40
2	O3464-02		1	14	100.00	>212.0	>212.0	06/29/2023	12:10
3	O3464-02DUP		1	13	100.00	>212.0	>212.0	06/29/2023	12:40
4	O3469-01		1	14	100.00	>212.0	>212.0	06/29/2023	13:10
5	O3469-02		1	14	100.00	>212.0	>212.0	06/29/2023	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6126252

WorkList Name : fp s 6-29

WorkList ID : 171462

Department : Wet-Chemistry

Date : 06-29-2023 08:18:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3464-02 A	365	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B
O3469-01A	2849-51-52-53	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B
O3469-02 B	2854	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B

Date/Time 06/29/2023 11:30
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: MCCLC

Date/Time 06/29/2023 14:15
 Raw Sample Received by: JAC
 Raw Sample Relinquished by: RM CWG