


LB126254

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.6927e-004 slope

C= -4.8220e-002 y-intercept

Rho= 0.9999597

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	0.027	0.027	445	0.000	445					
0.05	0.050					302					
0.20	0.200	0.185	-0.015	1375	0.0 %	1375	0				
2.50	2.500	2.489	-0.011	14990	0.0 %	14990					
5.00	5.000	4.964	-0.036	29611	0.0 %	29611					
7.50	7.500	7.560	0.060	44948	0.0 %	44948					
10.0	10.000	9.976	-0.024	59219	0.0 %	59219					

0.0017

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	445	-		0 SFAM01.1	PPB	6/29/2023 17:27	Std
	0.2 S01	1375	-		0.2 SFAM01.1	PPB	6/29/2023 17:29	Std
	2.5 S02	14990	-		2.5 SFAM01.1	PPB	6/29/2023 17:32	Std
	5 S03	29611	-		5 SFAM01.1	PPB	6/29/2023 17:34	Std
	7.5 S04	44948	-		7.5 SFAM01.1	PPB	6/29/2023 17:36	Std
	10 S05	59219	-		10 SFAM01.1	PPB	6/29/2023 17:39	Std
ICV049	ICV049	24392	4.0805 -		SFAM01.1	PPB	6/29/2023 17:44	SMPL
ICB049	ICB049	320	0.0059 -		SFAM01.1	PPB	6/29/2023 17:46	SMPL
CCV057	CCV057	29574	4.9576 -		SFAM01.1	PPB	6/29/2023 17:48	SMPL
CCB057	CCB057	318	0.0056 -		SFAM01.1	PPB	6/29/2023 17:50	SMPL
PB153906BL	PBS906	99	-0.0315 -		SFAM01.1	PPB	6/29/2023 17:53	SMPL
O3406-01	MJVA08	572	0.0486 -		SFAM01.1	PPB	6/29/2023 17:55	SMPL
O3406-02	MJVA09	985	0.1185 -		SFAM01.1	PPB	6/29/2023 17:57	SMPL
O3406-03	MJVA82	1146	0.1458 -		SFAM01.1	PPB	6/29/2023 17:59	SMPL
O3406-04	MJVA06	1136	0.1441 -		SFAM01.1	PPB	6/29/2023 18:02	SMPL
O3406-05	MJVA48	4367	0.691 -		SFAM01.1	PPB	6/29/2023 18:04	SMPL
O3406-06	MJVA50	1427	0.1933 -		SFAM01.1	PPB	6/29/2023 18:06	SMPL
O3406-07	MJVA52	1318	0.1749 -		SFAM01.1	PPB	6/29/2023 18:08	SMPL
O3406-08	MJVA53	1485	0.2031 -		SFAM01.1	PPB	6/29/2023 18:11	SMPL
O3406-09	MJVA63	2563	0.3856 -		SFAM01.1	PPB	6/29/2023 18:13	SMPL
O3406-10	MJVA64	1075	0.1337 -		SFAM01.1	PPB	6/29/2023 18:15	SMPL
O3406-11	MJVA73	1586	0.2202 -		SFAM01.1	PPB	6/29/2023 18:17	SMPL
O3406-12	MJVA73D	1559	0.2157 -		SFAM01.1	PPB	6/29/2023 18:20	SMPL
O3406-13	MJVA73S	16047	2.668 -		SFAM01.1	PPB	6/29/2023 18:22	SMPL
O3406-14	MJVA74	1525	0.2099 -		SFAM01.1	PPB	6/29/2023 18:24	SMPL
O3406-15	MJVA86	1008	0.1224 -		SFAM01.1	PPB	6/29/2023 18:26	SMPL
O3406-16	MJVA54	1305	0.1727 -		SFAM01.1	PPB	6/29/2023 18:29	SMPL
O3406-17	MJVA67	1037	0.1273 -		SFAM01.1	PPB	6/29/2023 18:31	SMPL
O3406-18	MJVA81	448	0.0276 -		SFAM01.1	PPB	6/29/2023 18:33	SMPL
O3406-19	MJVA96	1760	0.2497 -		SFAM01.1	PPB	6/29/2023 18:35	SMPL
CCV058	CCV058	30165	5.0577 -		SFAM01.1	PPB	6/29/2023 18:38	SMPL
CCB058	CCB058	339	0.0092 -		SFAM01.1	PPB	6/29/2023 18:40	SMPL
O3406-20	MJVA97	1588	0.2206 -		SFAM01.1	PPB	6/29/2023 18:42	SMPL
O3406-21	MJVAA6	1544	0.2131 -		SFAM01.1	PPB	6/29/2023 18:45	SMPL
O3406-22	MJVAA7	2414	0.3604 -		SFAM01.1	PPB	6/29/2023 18:47	SMPL
CCV059	CCV059	30277	5.0766 -		SFAM01.1	PPB	6/29/2023 18:49	SMPL
CCB059	CCB059	319	0.0058 -		SFAM01.1	PPB	6/29/2023 18:51	SMPL