

LB126264


7470A

INSTRUMENT 113: CV1

Linear

μ Abs.:

50961



A= 0.0000e+000

B= 1.9603e-004

C= -3.3882e-002

Rho= 0.9996854

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	468
0.0	0.000	0.014	0.014	242	0.000	242	0				-19
0.2	0.200	0.162	-0.038	1001	0.0 %	1001					-3
2.5	2.500	2.424	-0.076	12538	0.0 %	12538					-4
5.0	5.000	5.198	0.198	26689	0.0 %	26689					-1
7.5	7.500	7.446	-0.054	38158	0.0 %	38158					0
10.0	10.000	9.956	-0.044	50961	0.0 %	50961					

LB126264 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jun 2023 10:30:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	242	-	0.0000	7470A	PPB	30 Jun 2023 10:46:32	S	Std
0.2 - 1	50.2	1001	-	0.2000	7470A	PPB	30 Jun 2023 10:48:49	S	Std
2.5 - 1	52.5	12538	-	2.5000	7470A	PPB	30 Jun 2023 10:51:06	S	Std
5.0 - 1	55	26689	-	5.0000	7470A	PPB	30 Jun 2023 10:53:22	S	Std
7.5 - 1	57.5	38158	-	7.5000	7470A	PPB	30 Jun 2023 10:55:38	S	Std
10.0 - 1	510	50961	-	10.0000	7470A	PPB	30 Jun 2023 11:00:35	S	Std
ICV50 - 1	ICV50	20040	3.8946	-	7470A	PPB	30 Jun 2023 11:06:04	U	SMPL
ICB50 - 1	ICB50	-92	-0.0519	-	7470A	PPB	30 Jun 2023 11:08:18	U	SMPL
CCV60 - 1	CCV60	25104	4.8872	-	7470A	PPB	30 Jun 2023 11:10:35	U	SMPL
CCB60 - 1	CCB60	-105	-0.0545	-	7470A	PPB	30 Jun 2023 11:12:50	U	SMPL
CRA - 1	CRA	1105	0.1827	-	7470A	PPB	30 Jun 2023 11:15:06	U	SMPL
HighStd - 1	HighStd	51012	9.9660	-	7470A	PPB	30 Jun 2023 11:17:20	U	SMPL
ChkStd - 1	ChkStd	34393	6.7082	-	7470A	PPB	30 Jun 2023 11:19:35	U	SMPL
PB153901BL - 1	PBW	-231	-0.0792	-	7470A	PPB	30 Jun 2023 11:21:53	U	SMPL
PB153901BS - 1	LCSW	17239	3.3455	-	7470A	PPB	30 Jun 2023 11:24:11	U	SMPL
O3464-01 - 1	364	-133	-0.0600	-	7470A	PPB	30 Jun 2023 11:26:27	U	SMPL
O3464-02 - 1	365	19	-0.0302	-	7470A	PPB	30 Jun 2023 11:28:43	U	SMPL
O3465-01 - 1	3102	-117	-0.0568	-	7470A	PPB	30 Jun 2023 11:31:00	U	SMPL
O3465-02 - 1	3107-3109	98	-0.0147	-	7470A	PPB	30 Jun 2023 11:33:17	U	SMPL
O3469-01 - 1	2849-51-52-53	172	-0.0002	-	7470A	PPB	30 Jun 2023 11:35:32	U	SMPL
CCV61 - 1	CCV61	23880	4.6473	-	7470A	PPB	30 Jun 2023 11:37:47	U	SMPL
CCB61 - 1	CCB61	-100	-0.0535	-	7470A	PPB	30 Jun 2023 11:40:02	U	SMPL
O3469-02 - 1	2854	167	-0.0011	-	7470A	PPB	30 Jun 2023 11:42:19	U	SMPL
O3471-01 - 1	201	157	-0.0031	-	7470A	PPB	30 Jun 2023 11:44:34	U	SMPL
O3471-02 - 1	203	214	0.0081	-	7470A	PPB	30 Jun 2023 11:46:49	U	SMPL
O3471-02DUP - 1	203DUP	183	0.0020	-	7470A	PPB	30 Jun 2023 11:49:05	U	SMPL
O3471-02MS - 1	203MS	17241	3.3459	-	7470A	PPB	30 Jun 2023 11:51:20	U	SMPL
O3471-02MSD - 1	203MSD	19342	3.7577	-	7470A	PPB	30 Jun 2023 11:53:36	U	SMPL
PB153902BL - 1	PBW	-101	-0.0537	-	7470A	PPB	30 Jun 2023 11:55:54	U	SMPL
PB153902BS - 1	LCSW	18434	3.5797	-	7470A	PPB	30 Jun 2023 12:00:44	U	SMPL
O3464-03 - 1	366	-56	-0.0449	-	7470A	PPB	30 Jun 2023 12:03:00	U	SMPL
O3464-03DUP - 1	366DUP	52	-0.0237	-	7470A	PPB	30 Jun 2023 12:05:17	U	SMPL
CCV62 - 1	CCV62	24303	4.7302	-	7470A	PPB	30 Jun 2023 12:07:32	U	SMPL
CCB62 - 1	CCB62	-98	-0.0531	-	7470A	PPB	30 Jun 2023 12:09:47	U	SMPL
O3464-03MS - 1	366MS	18425	3.5780	-	7470A	PPB	30 Jun 2023 12:14:42	U	SMPL
O3464-03MSD - 1	366MSD	18020	3.4986	-	7470A	PPB	30 Jun 2023 12:16:57	U	SMPL
O3464-04 - 1	367	125	-0.0094	-	7470A	PPB	30 Jun 2023 12:19:14	U	SMPL
O3465-03 - 1	3110	214	0.0081	-	7470A	PPB	30 Jun 2023 12:21:31	U	SMPL
O3465-04 - 1	3041-3043	148	-0.0049	-	7470A	PPB	30 Jun 2023 12:23:46	U	SMPL
O3471-05 - 1	204	239	0.0130	-	7470A	PPB	30 Jun 2023 12:26:02	U	SMPL
PB153873TB - 1	PB153873TB	11	-0.0317	-	7470A	PPB	30 Jun 2023 12:28:18	U	SMPL
PB153874TB - 1	PB153874TB	68	-0.0206	-	7470A	PPB	30 Jun 2023 12:30:34	U	SMPL
CCV63 - 1	CCV63	24457	4.7604	-	7470A	PPB	30 Jun 2023 12:32:51	U	SMPL
CCB63 - 1	CCB63	19	-0.0302	-	7470A	PPB	30 Jun 2023 12:35:07	U	SMPL
O3471-02LX5 - 1		14	-0.0311	-	7470A	PPB	30 Jun 2023 12:37:24	U	SMPL
O3471-02A - 1		20868	4.0569	-	7470A	PPB	30 Jun 2023 12:39:40	U	SMPL
O3464-03LX5 - 1		-96	-0.0527	-	7470A	PPB	30 Jun 2023 12:41:55	U	SMPL
O3464-03A - 1		21046	4.0918	-	7470A	PPB	30 Jun 2023 12:44:12	U	SMPL
CCV64 - 1	CCV64	24894	4.8461	-	7470A	PPB	30 Jun 2023 12:46:27	U	SMPL
CCB64 - 1	CCB64	-9	-0.0356	-	7470A	PPB	30 Jun 2023 12:48:44	U	SMPL