



PERCENT SOLID

Supervisor: Sohil

Analyst: JIGNESH

Date: 7/12/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 13:40

In Date: 06/30/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:32

Out Date: 07/01/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB126266

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3446-01	MH0F02	1	1.00	1.00	2.00	2.00	100.0	
O3446-02	MH0F02D	2	1.00	1.00	2.00	2.00	100.0	
O3446-03	MH0F02S	3	1.00	1.00	2.00	2.00	100.0	
O3446-04	MH0F03	4	1.00	1.00	2.00	2.00	100.0	
O3446-05	MH0F07	5	1.00	1.00	2.00	2.00	100.0	
O3446-06	MH0F08	6	1.00	1.00	2.00	2.00	100.0	
O3446-07	MH0F09	7	1.00	1.00	2.00	2.00	100.0	
O3446-08	MH0F10	8	1.00	1.00	2.00	2.00	100.0	
O3446-09	MH0F11	9	1.00	1.00	2.00	2.00	100.0	
O3446-10	MH0F12	10	1.00	1.00	2.00	2.00	100.0	
O3446-11	MH0F13	11	1.00	1.00	2.00	2.00	100.0	
O3446-12	MH0F14	12	1.00	1.00	2.00	2.00	100.0	
O3446-13	MH0F27	13	1.00	1.00	2.00	2.00	100.0	
O3446-14	MH0F28	14	1.00	1.00	2.00	2.00	100.0	
O3446-15	MH0F29	15	1.00	1.00	2.00	2.00	100.0	
O3446-16	MH0F30	16	1.00	1.00	2.00	2.00	100.0	
O3446-17	MH0F31	17	1.00	1.00	2.00	2.00	100.0	
O3446-18	MH0F32	18	1.00	1.00	2.00	2.00	100.0	
O3446-19	MH0F33	19	1.00	1.00	2.00	2.00	100.0	
O3446-20	MH0F34	20	1.00	1.00	2.00	2.00	100.0	
O3446-21	MH0F35	21	1.00	1.00	2.00	2.00	100.0	
O3446-22	MH0F36	22	1.00	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$