



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126269

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/30/2023	08:15
2	O3464-03		1	14	100.00	>212.0	>212.0	06/30/2023	08:45
3	O3464-03DUP		1	14	100.00	>212.0	>212.0	06/30/2023	09:15
4	O3464-04		1	12	100.00	>212.0	>212.0	06/30/2023	09:45
5	O3465-03		1	13	100.00	>212.0	>212.0	06/30/2023	10:15
6	O3465-04		1	14	100.00	>212.0	>212.0	06/30/2023	10:45
7	O3471-05		1	12	100.00	>212.0	>212.0	06/30/2023	11:15
8	O3482-04		1	14	100.00	>212.0	>212.0	06/30/2023	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126269

WorkList Name : fp w-6-29

WorkList ID : 171463

Department : Wet-Chemistry

Date : 06-29-2023 08:19:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3464-03 C	366	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B
O3464-04 C	367	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B
O3465-03 C	3110	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B
O3465-04 C	3041-3043	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B
O3471-05 C	204	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/28/2023	1010B
O3482-04 F	COMP-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	I11	06/29/2023	1010B

Date/Time 06/30/2023 08:00
 Raw Sample Received by: Rm9 Cewe
 Raw Sample Relinquished by: JH6026

Date/Time 06/30/2023 12:30
 Raw Sample Received by: JH (CWC)
 Raw Sample Relinquished by: Rm9 (CWC)