



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/3/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 15:10

In Date: 06/30/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:10

Out Date: 07/01/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: % SOLIDS-OVEN

QC:LB126276

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3439-01	A46A6	1	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3439-02	A46C1	2	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3439-03	A46K7	3	1.13	8.67	9.8	8.22	81.8	
O3439-04	A46K7D	4	1.13	8.67	9.8	8.22	81.8	
O3439-05	A46K7S	5	1.13	8.67	9.8	8.22	81.8	
O3439-06	A4630	6	1.12	8.82	9.94	6.76	63.9	
O3439-07	A4631	7	1.16	8.59	9.75	9.00	91.3	
O3439-08	A46H4	8	1.17	8.80	9.97	3.93	31.4	
O3439-09	A46H5	9	1.19	8.56	9.75	8.18	81.7	
O3439-10	A46H6	10	1.18	8.49	9.67	8.75	89.2	
O3439-11	A46H7	11	1.18	8.75	9.93	5.49	49.3	
O3439-12	A46H8	12	1.15	8.83	9.98	9.47	94.2	
O3439-13	A46G6	13	1.12	8.84	9.96	4.4	37.1	
O3439-14	A46G7	14	1.11	8.47	9.58	3.66	30.1	
O3439-15	A46G8	15	1.14	8.79	9.93	9.25	92.3	
O3439-16	A46G9	16	1.18	8.81	9.99	9.1	89.9	
O3439-17	A46H1	17	1.14	8.78	9.92	8.66	85.6	
O3439-18	A46H2	18	1.15	8.81	9.96	9.21	91.5	
O3439-19	A46K9	19	1.15	8.79	9.94	8.99	89.2	
O3439-20	A46L0	20	1.14	8.83	9.97	8.92	88.1	
O3439-21	A46L7	21	1.18	8.67	9.85	3.27	24.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$