



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/3/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 16:30

In Date: 06/30/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 07/01/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB126280

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3492-01	TB-TP1	1	1.19	8.42	9.61	8.44	86.1	
O3492-03	TB-TP1-EPH-2	2	1.18	8.51	9.69	8.7	88.4	
O3492-04	TB-TP2	3	1.11	8.50	9.61	8.54	87.4	
O3492-06	TB-TP2-EPH-2	4	1.17	8.57	9.74	8.9	90.2	
O3492-07	TB-CONC	5	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3493-01	ACM-CAULKING	6	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3494-05	SVOC-GPC-BLANK	7	1.00	1.00	2.00	2.00	100.0	
O3494-06	PEST-GPC-BLANK	8	1.00	1.00	2.00	2.00	100.0	
O3494-07	PEST-GPC-BLANK-SPIKE	9	1.00	1.00	2.00	2.00	100.0	
O3494-08	PCB-GPC-BLANK	10	1.00	1.00	2.00	2.00	100.0	
O3494-09	PCB-GPC-BLANK-SPIKE	11	1.00	1.00	2.00	2.00	100.0	
O3494-10	SVOC-GPC2-BLANK	12	1.00	1.00	2.00	2.00	100.0	
O3494-11	PEST-GPC2-BLANK	13	1.00	1.00	2.00	2.00	100.0	
O3494-12	PEST-GPC2-BLANK-SPIKE	14	1.00	1.00	2.00	2.00	100.0	
O3494-13	PCB-GCP2-BLANK	15	1.00	1.00	2.00	2.00	100.0	
O3494-14	PCB-GCP2-BLANK-SPIKE	16	1.00	1.00	2.00	2.00	100.0	
O3496-01	01A-01B-01C	17	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3496-02	02A-02B-02C	18	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3496-03	03A-03B-03C	19	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3496-04	04A-04B-04C	20	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3497-01	OK-2-063023	21	1.18	8.63	9.81	9.03	91.0	
O3497-02	OK-2-063023-EPH-2	22	1.16	8.65	9.81	9.00	90.6	
O3498-01	NB-07-063023	23	1.19	8.53	9.72	9.21	94.0	
O3498-02	NB-07-063023-EPH-2	24	1.19	8.53	9.72	9.21	94.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$