



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/3/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 17:35

In Date: 06/30/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:20

Out Date: 07/01/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB126286

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3452-01	A46G4	1	1.18	8.49	9.67	8.8	89.8	
O3452-02	A46G5	2	1.12	8.86	9.98	9.51	94.7	
O3452-03	VHBLK001	3	1.00	1.00	2.00	2.00	100.0	VHBLK
O3452-04	A4665	4	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3452-05	A4681	5	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3452-06	A46N3	6	1.12	8.50	9.62	8.81	90.5	
O3452-07	A46N4	7	1.15	8.82	9.97	7.33	70.1	
O3452-08	A46N5	8	1.14	8.78	9.92	6.63	62.5	
O3452-09	A46N6	9	1.19	8.37	9.56	8.71	89.8	
O3452-10	A46N8	10	1.17	8.80	9.97	3.94	31.5	
O3452-11	A46N9	11	1.15	8.82	9.97	5.69	51.5	
O3452-12	A46P0	12	1.17	8.75	9.92	8.87	88.0	
O3452-13	A46P3	13	1.19	8.62	9.81	6.52	61.8	
O3452-14	A46P4	14	1.11	8.86	9.97	7.72	74.6	
O3452-15	A46P7	15	1.17	8.58	9.75	4.07	33.8	
O3452-16	A46P8	16	1.14	8.82	9.96	4.97	43.4	
O3452-17	A46P9	17	1.16	8.61	9.77	8.91	90.0	
O3452-18	A46Q6	18	1.15	8.84	9.99	3.05	21.5	
O3452-19	A46Q7	19	1.12	8.83	9.95	5.79	52.9	
O3452-20	A46Q8	20	1.13	8.71	9.84	8.8	88.1	
O3452-21	A46Q0	21	1.16	8.81	9.97	3.51	26.7	
O3452-22	A46Q0MS	22	1.16	8.81	9.97	3.51	26.7	
O3452-23	A46Q0MSD	23	1.16	8.81	9.97	3.51	26.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$