

LB126290


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53447



A= 0.0000e+000

B= 1.8916e-004 *slope*C= -5.1475e-003 *+ intercept*

Rho= 0.9998200

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	0.054	0.054	314	0.000	314					
0.05	0.050					302	0				
0.20	0.200	0.219	0.019	1186	0.0 %	1186					
2.50	2.500	2.462	-0.038	13041	0.0 %	13041					
5.00	5.000	4.974	-0.026	26325	0.0 %	26325					
7.50	7.500	7.386	-0.114	39074	0.0 %	39074					
10.0	10.000	10.105	0.105	53447	0.0 %	53447					

10
-2
-1
-2
1

LB126290

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	314	-		0 SFAM01.1	PPB	6/30/2023 17:47	Std
	0.2 S01	1186	-		0.2 SFAM01.1	PPB	6/30/2023 17:49	Std
	2.5 S02	13041	-		2.5 SFAM01.1	PPB	6/30/2023 17:51	Std
	5 S03	26325	-		5 SFAM01.1	PPB	6/30/2023 17:54	Std
	7.5 S04	39074	-		7.5 SFAM01.1	PPB	6/30/2023 17:56	Std
	10 S05	53447	-		10 SFAM01.1	PPB	6/30/2023 18:01	Std
ICV052	ICV052	19999	3.7778 -		SFAM01.1	PPB	6/30/2023 18:06	SMPL
ICB052	ICB052	187	0.0302 -		SFAM01.1	PPB	6/30/2023 18:08	SMPL
CCV070	CCV070	26296	4.9689 -		SFAM01.1	PPB	6/30/2023 18:11	SMPL
CCB070	CCB070	185	0.0298 -		SFAM01.1	PPB	6/30/2023 18:13	SMPL
PB153939BL	PBW939	304	0.0524 -		SFAM01.1	PPB	6/30/2023 18:15	SMPL
O3292-03	MH0F91	350	0.0611 -		SFAM01.1	PPB	6/30/2023 18:17	SMPL
O3292-04	MH0F92	345	0.0601 -		SFAM01.1	PPB	6/30/2023 18:20	SMPL
O3292-05	MH0F93	355	0.062 -		SFAM01.1	PPB	6/30/2023 18:22	SMPL
O3292-06	MH0F93D	370	0.0648 -		SFAM01.1	PPB	6/30/2023 18:24	SMPL
O3292-07	MH0F93S	5238	0.9856 -		SFAM01.1	PPB	6/30/2023 18:26	SMPL
O3292-08	MH0F94	467	0.0832 -		SFAM01.1	PPB	6/30/2023 18:29	SMPL
O3292-09	MH0F95	431	0.0764 -		SFAM01.1	PPB	6/30/2023 18:31	SMPL
O3292-10	MH0F89	357	0.0624 -		SFAM01.1	PPB	6/30/2023 18:33	SMPL
O3292-11	MH0F90	383	0.0673 -		SFAM01.1	PPB	6/30/2023 18:35	SMPL
CCV071	CCV071	25912	4.8962 -		SFAM01.1	PPB	6/30/2023 18:38	SMPL
CCB071	CCB071	165	0.0261 -		SFAM01.1	PPB	6/30/2023 18:40	SMPL