

LB126296


7470A

INSTRUMENT ID: EVL

Linear

 μ Abs.:

48641



A= 0.0000e+000

B= 2.0368e-004

C= -1.5869e-003

Rho= 0.9997579

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	A/D
0.0	0.000	0.047	0.047	237	0.000	237	0				-19
0.2	0.200	0.162	-0.038	805	0.0 %	805					-4
2.5	2.500	2.400	-0.100	11792	0.0 %	11792					2
5.0	5.000	5.091	0.091	25002	0.0 %	25002					1
7.5	7.500	7.594	0.094	37292	0.0 %	37292					-1
10.0	10.000	9.906	-0.094	48641	0.0 %	48641					-1

LB126296 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jul 2023 13:49:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	237	-	0.0000	7470A	PPB	03 Jul 2023 14:00:59	S	Std
0.2 - 1	50-2	805	-	0.2000	7470A	PPB	03 Jul 2023 14:03:15	S	Std
2.5 - 1	52.5	11792	-	2.5000	7470A	PPB	03 Jul 2023 14:05:32	S	Std
5.0 - 1	55	25002	-	5.0000	7470A	PPB	03 Jul 2023 14:07:49	S	Std
7.5 - 1	57.5	37292	-	7.5000	7470A	PPB	03 Jul 2023 14:10:05	S	Std
10.0 - 1	510	48641	-	10.0000	7470A	PPB	03 Jul 2023 14:20:20	S	Std
ICV53 - 1	ICV53	20378	4.1490	-	7470A	PPB	03 Jul 2023 14:23:06	U	SMPL
ICB53 - 1	ICB53	-28	-0.0073	-	7470A	PPB	03 Jul 2023 14:25:20	U	SMPL
CCV72 - 1	CCV72	23682	4.8220	-	7470A	PPB	03 Jul 2023 14:27:36	U	SMPL
CCB72 - 1	CCB72	-35	-0.0087	-	7470A	PPB	03 Jul 2023 14:29:51	U	SMPL
CRA - 1	CRA	911	0.1840	-	7470A	PPB	03 Jul 2023 14:32:07	U	SMPL
HighStd - 1	HighStd	48284	9.8330	-	7470A	PPB	03 Jul 2023 14:34:22	U	SMPL
ChkStd - 1	ChkStd	32535	6.6252	-	7470A	PPB	03 Jul 2023 14:36:37	U	SMPL
PB153950BL - 1	PBW	48	0.0082	-	7470A	PPB	03 Jul 2023 14:38:55	U	SMPL
PB153950BS - 1	LCSW	17655	3.5944	-	7470A	PPB	03 Jul 2023 14:41:12	U	SMPL
O3368-07 - 1	CF-28	143	0.0275	-	7470A	PPB	03 Jul 2023 14:43:28	U	SMPL
O3368-07DUP - 1	CF-28DUP	342	0.0681	-	7470A	PPB	03 Jul 2023 14:45:45	U	SMPL
O3368-07MS - 1	CF-28MS	19866	4.0448	-	7470A	PPB	03 Jul 2023 14:50:34	U	SMPL
O3368-07MSD - 1	CF-28MSD	18387	3.7435	-	7470A	PPB	03 Jul 2023 14:52:50	U	SMPL
O3368-08 - 1	CF-29	1496	0.3031	-	7470A	PPB	03 Jul 2023 14:55:06	U	SMPL
CCV73 - 1	CCV73	23575	4.8002	-	7470A	PPB	03 Jul 2023 14:57:22	U	SMPL
CCB73 - 1	CCB73	-66	-0.0150	-	7470A	PPB	03 Jul 2023 14:59:38	U	SMPL
PB153951BL - 1	PBW	250	0.0493	-	7470A	PPB	03 Jul 2023 15:01:54	U	SMPL
PB153951BS - 1	LCSW	19554	3.9812	-	7470A	PPB	03 Jul 2023 15:06:44	U	SMPL
O3403-04 - 1	ORA-1652	133	0.0255	-	7470A	PPB	03 Jul 2023 15:08:59	U	SMPL
O3482-01 - 1	COMP-1	297	0.0589	-	7470A	PPB	03 Jul 2023 15:11:15	U	SMPL
O3482-02 - 1	COMP-2	301	0.0597	-	7470A	PPB	03 Jul 2023 15:13:33	U	SMPL
O3482-03 - 1	COMP-3	253	0.0499	-	7470A	PPB	03 Jul 2023 15:15:51	U	SMPL
O3488-02 - 1	WC-1-COMP	303	0.0601	-	7470A	PPB	03 Jul 2023 15:18:07	U	SMPL
O3488-02DUP - 1	WC-1-COMPDUP	424	0.0848	-	7470A	PPB	03 Jul 2023 15:20:24	U	SMPL
O3488-02MS - 1	WC-1-COMPMS	19595	3.9896	-	7470A	PPB	03 Jul 2023 15:22:41	U	SMPL
O3488-02MSD - 1	WC-1-COMPMSD	19381	3.9460	-	7470A	PPB	03 Jul 2023 15:24:56	U	SMPL
CCV74 - 1	CCV74	23936	4.8737	-	7470A	PPB	03 Jul 2023 15:27:12	U	SMPL
CCB74 - 1	CCB74	28	0.0041	-	7470A	PPB	03 Jul 2023 15:29:28	U	SMPL
PB153952BL - 1	PBW	236	0.0465	-	7470A	PPB	03 Jul 2023 15:31:45	U	SMPL
PB153952BS - 1	LCSW	16948	3.4504	-	7470A	PPB	03 Jul 2023 15:34:00	U	SMPL
O3492-02 - 1	TB-TP1	146	0.0282	-	7470A	PPB	03 Jul 2023 15:36:16	U	SMPL
O3492-02DUP - 1	TB-TP1DUP	440	0.0880	-	7470A	PPB	03 Jul 2023 15:38:32	U	SMPL
O3492-02MS - 1	TB-TP1MS	19222	3.9136	-	7470A	PPB	03 Jul 2023 15:51:31	U	SMPL
O3492-02MSD - 1	TB-TP1MSD	19138	3.8965	-	7470A	PPB	03 Jul 2023 15:53:47	U	SMPL
O3492-05 - 1	TB-TP2	35	0.0055	-	7470A	PPB	03 Jul 2023 15:56:04	U	SMPL
PB153953BL - 1	PBW	36	0.0057	-	7470A	PPB	03 Jul 2023 15:58:22	U	SMPL
PB153953BS - 1	LCSW	17582	3.5796	-	7470A	PPB	03 Jul 2023 16:00:39	U	SMPL
O3482-04 - 1	COMP-TOTE	174	0.0339	-	7470A	PPB	03 Jul 2023 16:02:54	U	SMPL
CCV75 - 1	CCV75	23851	4.8564	-	7470A	PPB	03 Jul 2023 16:05:10	U	SMPL
CCB75 - 1	CCB75	48	0.0082	-	7470A	PPB	03 Jul 2023 16:07:26	U	SMPL
O3482-04DUP - 1	COMP-TOTEDUP	278	0.0550	-	7470A	PPB	03 Jul 2023 16:09:42	U	SMPL
O3482-04MS - 1	COMP-TOTEMS	16845	3.4294	-	7470A	PPB	03 Jul 2023 16:11:57	U	SMPL
O3482-04MSD - 1	COMP-TOTEMSD	18519	3.7704	-	7470A	PPB	03 Jul 2023 16:14:12	U	SMPL
PB153950TB - 1	PB153898TB	108	0.0204	-	7470A	PPB	03 Jul 2023 16:16:28	U	SMPL
PB153899TB - 1	PB153899TB	213	0.0418	-	7470A	PPB	03 Jul 2023 16:18:45	U	SMPL
PB153930TB - 1	PB153930TB	250	0.0493	-	7470A	PPB	03 Jul 2023 16:21:01	U	SMPL
PB153932TB - 1	PB153932TB	274	0.0542	-	7470A	PPB	03 Jul 2023 16:23:17	U	SMPL
CCV76 - 1	CCV76	24487	4.9860	-	7470A	PPB	03 Jul 2023 16:25:34	U	SMPL
CCB76 - 1	CCB76	36	0.0057	-	7470A	PPB	03 Jul 2023 16:27:51	U	SMPL
O3368-07LX5 - 1		86	0.0159	-	7470A	PPB	03 Jul 2023 16:30:07	U	SMPL
O3368-07A - 1		15554	3.1665	-	7470A	PPB	03 Jul 2023 16:32:23	U	SMPL
O3488-02LX5 - 1		-103	-0.0226	-	7470A	PPB	03 Jul 2023 16:34:38	U	SMPL
O3488-02A - 1		17909	3.6462	-	7470A	PPB	03 Jul 2023 16:36:55	U	SMPL
O3492-02LX5 - 1		16	0.0017	-	7470A	PPB	03 Jul 2023 16:39:10	U	SMPL
O3492-02A - 1		20411	4.1558	-	7470A	PPB	03 Jul 2023 16:41:27	U	SMPL
O3482-04LX5 - 1		2	-0.0012	-	7470A	PPB	03 Jul 2023 16:43:42	U	SMPL
O3482-04A - 1		16722	3.4044	-	7470A	PPB	03 Jul 2023 16:45:58	U	SMPL
CCV77 - 1	CCV77	25053	5.1013	-	7470A	PPB	03 Jul 2023 16:48:14	U	SMPL
CCB77 - 1	CCB77	125	0.0239	-	7470A	PPB	03 Jul 2023 16:50:31	U	SMPL