

LB126298


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

55759



A= 0.0000e+000

B= 1.7853e-004 *slope*

C= -1.8852e-003 *y-intercept*

Rho= 0.9998978

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/b</i>
0.00	0.000	0.023	0.023	142	0.000	142					
0.05	0.050					302	0				
0.20	0.200	0.145	-0.055	822	0.0 %	822					
2.50	2.500	2.474	-0.026	13871	0.0 %	13871					
5.00	5.000	5.101	0.101	28583	0.0 %	28583					
7.50	7.500	7.504	0.004	42041	0.0 %	42041					
10.0	10.000	9.953	-0.047	55759	0.0 %	55759					

0/0
-28
000100

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	142	-		0 SFAM01.1	PPB	7/3/2023 17:49	Std
	0.2 S01	822	-		0.2 SFAM01.1	PPB	7/3/2023 17:51	Std
	2.5 S02	13871	-		2.5 SFAM01.1	PPB	7/3/2023 17:53	Std
	5 S03	28583	-		5 SFAM01.1	PPB	7/3/2023 17:55	Std
	7.5 S04	42041	-		7.5 SFAM01.1	PPB	7/3/2023 17:58	Std
	10 S05	55759	-		10 SFAM01.1	PPB	7/3/2023 18:03	Std
ICV054	ICV054	22224	3.9657 -		SFAM01.1	PPB	7/3/2023 18:08	SMPL
ICB054	ICB054	35	0.0044 -		SFAM01.1	PPB	7/3/2023 18:10	SMPL
CCV078	CCV078	28540	5.0933 -		SFAM01.1	PPB	7/3/2023 18:12	SMPL
CCB078	CCB078	5	-0.001 -		SFAM01.1	PPB	7/3/2023 18:15	SMPL
PB153980BL	PBS980	175	0.0294 -		SFAM01.1	PPB	7/3/2023 18:17	SMPL
O3426-02	A46F3	25207	4.4983 -		SFAM01.1	PPB	7/3/2023 18:19	SMPL
O3426-03	A46F4	269	0.0461 -		SFAM01.1	PPB	7/3/2023 18:21	SMPL
O3426-04	A4633	17699	3.1579 -		SFAM01.1	PPB	7/3/2023 18:24	SMPL
O3426-05	A4635	2111	0.375 -		SFAM01.1	PPB	7/3/2023 18:26	SMPL
O3426-06	A4642	3885	0.6917 -		SFAM01.1	PPB	7/3/2023 18:28	SMPL
O3426-07	A4643	20309	3.6238 -		SFAM01.1	PPB	7/3/2023 18:31	SMPL
O3426-08	A4644	33521	5.9825 -		SFAM01.1	PPB	7/3/2023 18:33	SMPL
O3426-09	A46E9	9936	1.772 -		SFAM01.1	PPB	7/3/2023 18:35	SMPL
O3426-10	A46F0	8100	1.4442 -		SFAM01.1	PPB	7/3/2023 18:37	SMPL
O3426-11	A46F1	280	0.0481 -		SFAM01.1	PPB	7/3/2023 18:40	SMPL
O3426-13	A46J1	19996	3.568 -		SFAM01.1	PPB	7/3/2023 18:42	SMPL
O3426-14	A46J2	6029	1.0745 -		SFAM01.1	PPB	7/3/2023 18:44	SMPL
O3426-15	A46J3	1847	0.3279 -		SFAM01.1	PPB	7/3/2023 18:46	SMPL
O3426-16	A46J4	6116	1.09 -		SFAM01.1	PPB	7/3/2023 18:49	SMPL
O3426-17	A46J6	3052	0.543 -		SFAM01.1	PPB	7/3/2023 18:51	SMPL
O3426-18	A46J7	1230	0.2177 -		SFAM01.1	PPB	7/3/2023 18:53	SMPL
O3426-19	A46J8	510	0.0892 -		SFAM01.1	PPB	7/3/2023 18:56	SMPL
O3426-20	A46J9	562	0.0984 -		SFAM01.1	PPB	7/3/2023 18:58	SMPL
O3426-21	A46K0	1183	0.2093 -		SFAM01.1	PPB	7/3/2023 19:00	SMPL
CCV079	CCV079	28820	5.1433 -		SFAM01.1	PPB	7/3/2023 19:02	SMPL
CCB079	CCB079	24	0.0024 -		SFAM01.1	PPB	7/3/2023 19:05	SMPL
O3426-22	A46K6	19162	3.4191 -		SFAM01.1	PPB	7/3/2023 19:07	SMPL
O3426-23	A46K6D	18105	3.2304 -		SFAM01.1	PPB	7/3/2023 19:09	SMPL
O3426-24	A46K6S	32553	5.8097 -		SFAM01.1	PPB	7/3/2023 19:11	SMPL
CCV080	CCV080	28673	5.117 -		SFAM01.1	PPB	7/3/2023 19:14	SMPL
CCB080	CCB080	-20	-0.0055 -		SFAM01.1	PPB	7/3/2023 19:16	SMPL