



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 7/6/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 13:50

In Date: 07/05/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:55

Out Date: 07/06/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: % SOLIDS-OVEN

QC:LB126308

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3489-01	A46N7	1	1.13	8.52	9.65	8.59	87.6	
O3489-02	A46P1	2	1.14	8.82	9.96	8.34	81.6	
O3489-03	A46P2	3	1.15	8.40	9.55	7.58	76.5	
O3489-04	A46P5	4	1.12	8.82	9.94	8.08	78.9	
O3489-05	A46R2	5	1.15	8.80	9.95	7.22	69.0	
O3489-06	A46R3	6	1.16	8.61	9.77	4.45	38.2	
O3489-07	A46R4	7	1.11	8.63	9.74	6.11	57.9	
O3489-08	A46R5	8	1.14	8.82	9.96	8.28	81.0	
O3489-09	A46R6	9	1.14	8.68	9.82	9.14	92.2	
O3489-10	A46Q3	10	1.17	8.81	9.98	6.5	60.5	
O3489-11	A46Q4	11	1.15	8.45	9.6	8.55	87.6	
O3489-12	A46Q5	12	1.16	8.49	9.65	7.93	79.7	
O3489-13	A46Q9	13	1.15	8.82	9.97	3.47	26.3	
O3489-14	A46R0	14	1.16	8.83	9.99	6.3	58.2	
O3489-15	A46R1	15	1.12	8.56	9.68	8.92	91.1	
O3489-16	A46Q1	16	1.12	8.65	9.77	6.06	57.1	
O3489-17	A46Q1D	17	1.12	8.65	9.77	6.06	57.1	
O3489-18	A46Q1S	18	1.12	8.65	9.77	6.06	57.1	
O3489-19	A46A8	19	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3489-20	A46B1	20	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3489-21	A46R8	21	1.14	8.83	9.97	8.1	78.8	
O3489-22	A46S4	22	1.16	8.80	9.96	4.41	36.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$