



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 7/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 07/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 07/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB126309

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3490-02	A46P1	1	1.14	8.82	9.96	8.34	81.6	
O3490-03	A46P2	2	1.15	8.40	9.55	7.58	76.5	
O3490-05	A46R2	3	1.15	8.80	9.95	7.22	69.0	
O3490-06	A46R3	4	1.16	8.61	9.77	4.45	38.2	
O3490-07	A46R4	5	1.11	8.63	9.74	6.11	57.9	
O3490-08	A46R5	6	1.14	8.82	9.96	8.28	81.0	
O3490-10	A46Q3	7	1.17	8.81	9.98	6.5	60.5	
O3490-11	A46Q4	8	1.15	8.45	9.6	8.55	87.6	
O3490-12	A46Q5	9	1.16	8.49	9.65	7.93	79.7	
O3490-13	A46Q9	10	1.15	8.82	9.97	3.47	26.3	
O3490-14	A46R0	11	1.16	8.83	9.99	6.3	58.2	
O3490-15	A46R1	12	1.12	8.56	9.68	8.92	91.1	
O3490-16	A46Q2	13	1.17	8.80	9.97	8.5	83.3	
O3490-17	A46Q2MS	14	1.17	8.80	9.97	8.5	83.3	
O3490-18	A46Q2MSD	15	1.17	8.80	9.97	8.5	83.3	
O3490-19	A46S7	16	1.12	8.67	9.79	8.00	79.4	
O3490-20	A46T0	17	1.19	8.71	9.9	8.68	86.0	
O3490-21	A46S4	18	1.16	8.80	9.96	4.41	36.9	
O3490-22	A46S5	19	1.11	8.47	9.58	7.55	76.0	
O3490-23	A46S9	20	1.13	8.84	9.97	5.8	52.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$