



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 7/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:55
In Date: 07/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 07/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB126310

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3444-01	A4672	1	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3444-02	A4678	2	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3444-03	A46N7	3	1.13	8.52	9.65	8.59	87.6	
O3444-04	A46P5	4	1.12	8.82	9.94	8.08	78.9	
O3444-05	A46R6	5	1.14	8.68	9.82	9.14	92.2	
O3444-06	A46T1	6	1.00	1.00	2.00	2.00	100.0	F.B.
O3444-07	A46Q1	7	1.12	8.65	9.77	6.06	57.1	
O3444-08	A46Q1MS	8	1.12	8.65	9.77	6.06	57.1	
O3444-09	A46Q1MSD	9	1.12	8.65	9.77	6.06	57.1	
O3444-10	VHBLK001	10	1.00	1.00	2.00	2.00	100.0	VHBLK
O3444-11	A46R8	11	1.14	8.83	9.97	8.1	78.8	
O3444-12	A46T6	12	1.14	8.81	9.95	8.22	80.4	
O3444-13	A46T7	13	1.12	8.63	9.75	6.08	57.5	
O3444-14	A46T8	14	1.11	8.44	9.55	8.42	86.6	
O3444-15	A46W0	15	1.12	8.40	9.52	4.82	44.0	
O3444-16	A46X0	16	1.00	1.00	2.00	2.00	100.0	F.B.
O3444-17	A46X8	17	1.13	8.84	9.97	6.11	56.3	
O3444-19	A46Y0	18	1.14	8.40	9.54	4.48	39.8	
O3444-20	A46Y1	19	1.13	8.85	9.98	5.96	54.6	
O3444-21	A46Y2	20	1.13	8.77	9.9	4.65	40.1	
O3444-22	A46Y3	21	1.12	8.79	9.91	8.79	87.3	
O3444-23	A4717	22	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
O3444-24	A4657	23	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$