

LB126311


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

6509E



A= 0.0000e+000

B= 1.5387e-004

C= -1.0608e-001

Rho= 0.9997723

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.086	-0.086	130	0.000	130					1
0.2	0.200	0.170	-0.030	1797	0.0 %	1797					1
2.5	2.500	2.573	0.073	17408	0.0 %	17408					1
5.0	5.000	5.125	0.125	33994	0.0 %	33994					3
7.5	7.500	7.508	0.008	49484	0.0 %	49484					3
10.0	10.000	9.910	-0.090	65095	0.0 %	65095					1

LB126311 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jul 2023 12:54:16

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	130	-	0.00007471B	PPB	05 Jul 2023 13:08:37	S	Std
0.2 - 1	50.2	1797	-	0.20007471B	PPB	05 Jul 2023 13:10:53	S	Std
2.5 - 1	52.5	17408	-	2.50007471B	PPB	05 Jul 2023 13:13:10	S	Std
5.0 - 1	55	33994	-	5.00007471B	PPB	05 Jul 2023 13:15:26	S	Std
7.5 - 1	57.5	49484	-	7.50007471B	PPB	05 Jul 2023 13:20:17	S	Std
10.0 - 1	510	65095	-	10.00007471B	PPB	05 Jul 2023 13:25:13	S	Std
ICV55 - 1	ICV55	27605	4.1416	-7471B	PPB	05 Jul 2023 13:28:23	U	SMPL
ICB55 - 1	ICB55	199	-0.0755	-7471B	PPB	05 Jul 2023 13:33:10	U	SMPL
CCV81 - 1	CCV81	33851	5.1027	-7471B	PPB	05 Jul 2023 13:35:24	U	SMPL
CCB81 - 1	CCB81	243	-0.0687	-7471B	PPB	05 Jul 2023 13:40:12	U	SMPL
CRA - 1	CRA	1816	0.1734	-7471B	PPB	05 Jul 2023 13:42:27	U	SMPL
HighStd - 1	HighStd	65870	10.0295	-7471B	PPB	05 Jul 2023 13:44:41	U	SMPL
ChkStd - 1	ChkStd	47022	7.1293	-7471B	PPB	05 Jul 2023 13:46:56	U	SMPL
PB153989BL - 1	PBS	258	-0.0664	-7471B	PPB	05 Jul 2023 13:52:38	U	SMPL
PB153989BS - 1	LCSS	27924	4.1907	-7471B	PPB	05 Jul 2023 13:54:54	U	SMPL
O3492-01 - 1	TB-TP1	78697	12.0033	-7471B	PPB	05 Jul 2023 13:57:10	U	SMPL
O3492-04 - 1	TB-TP2	102345	15.6420	-7471B	PPB	05 Jul 2023 13:59:28	U	SMPL
O3497-01 - 1	OK-2-063023	767	0.0119	-7471B	PPB	05 Jul 2023 14:01:47	U	SMPL
O3498-01 - 1	NB-07-063023	4826	0.6365	-7471B	PPB	05 Jul 2023 14:04:07	U	SMPL
O3507-01 - 1	CP-1	2719	0.3123	-7471B	PPB	05 Jul 2023 14:06:22	U	SMPL
CCV82 - 1	CCV82	33646	5.0711	-7471B	PPB	05 Jul 2023 14:08:38	U	SMPL
CCB82 - 1	CCB82	79	-0.0939	-7471B	PPB	05 Jul 2023 14:10:53	U	SMPL
O3507-01DUP - 1	CP-1DUP	2982	0.3528	-7471B	PPB	05 Jul 2023 14:13:11	U	SMPL
O3507-01MS - 1	CP-1MS	29897	4.4943	-7471B	PPB	05 Jul 2023 14:15:26	U	SMPL
O3507-01MSD - 1	CP-1MSD	28968	4.3513	-7471B	PPB	05 Jul 2023 14:17:41	U	SMPL
O3507-04 - 1	CP-2	2544	0.2854	-7471B	PPB	05 Jul 2023 14:19:58	U	SMPL
O3507-07 - 1	CP-3	5506	0.7411	-7471B	PPB	05 Jul 2023 14:22:17	U	SMPL
O3509-01 - 1	WC-WWS-01	115106	17.6056	-7471B	PPB	05 Jul 2023 14:24:35	U	SMPL
O3509-06 - 1	WC-WWS-02	94577	14.4468	-7471B	PPB	05 Jul 2023 14:26:51	U	SMPL
O3509-11 - 1	WC-WWS-03	127201	19.4667	-7471B	PPB	05 Jul 2023 14:29:18	U	SMPL
O3509-16 - 1	WC-WWS-04	81552	12.4426	-7471B	PPB	05 Jul 2023 14:31:38	U	SMPL
O3509-21 - 1	WC-WWS-05	90132	13.7628	-7471B	PPB	05 Jul 2023 14:33:57	U	SMPL
CCV83 - 1	CCV83	33282	5.0151	-7471B	PPB	05 Jul 2023 14:36:15	U	SMPL
CCB83 - 1	CCB83	113	-0.0887	-7471B	PPB	05 Jul 2023 14:42:47	U	SMPL
O3510-01 - 1	WC-WWS-06	31365	4.7201	-7471B	PPB	05 Jul 2023 14:45:02	U	SMPL
O3510-06 - 1	WC-WWS-07	38999	5.8948	-7471B	PPB	05 Jul 2023 14:47:18	U	SMPL
O3510-11 - 1	WC-WWS-10	67565	10.2903	-7471B	PPB	05 Jul 2023 14:49:35	U	SMPL
O3510-16 - 1	WC-WWS-11	130847	20.0277	-7471B	PPB	05 Jul 2023 14:51:52	U	SMPL
O3513-01 - 1	500-1-AB	169	-0.0801	-7471B	PPB	05 Jul 2023 14:54:10	U	SMPL
O3507-01LX5 - 1		262	-0.0658	-7471B	PPB	05 Jul 2023 14:56:36	U	SMPL
O3507-01A - 1		27713	4.1582	-7471B	PPB	05 Jul 2023 14:58:52	U	SMPL
CCV84 - 1	CCV84	32989	4.9700	-7471B	PPB	05 Jul 2023 15:01:09	U	SMPL
CCB84 - 1	CCB84	117	-0.0881	-7471B	PPB	05 Jul 2023 15:03:26	U	SMPL
O3492-01DLX5 - 1		16999	2.5096	-7471B	PPB	05 Jul 2023 15:05:43	U	SMPL
O3492-04DLX5 - 1		21998	3.2788	-7471B	PPB	05 Jul 2023 15:07:58	U	SMPL
O3509-01DLX5 - 1		24915	3.7277	-7471B	PPB	05 Jul 2023 15:10:15	U	SMPL
O3509-06DLX5 - 1		20006	2.9723	-7471B	PPB	05 Jul 2023 15:12:31	U	SMPL
O3509-11DLX5 - 1		21397	3.1863	-7471B	PPB	05 Jul 2023 15:14:47	U	SMPL
O3509-16DLX5 - 1		17975	2.6598	-7471B	PPB	05 Jul 2023 15:17:03	U	SMPL
O3509-21DLX5 - 1		19574	2.9058	-7471B	PPB	05 Jul 2023 15:19:21	U	SMPL
O3510-11DLX5 - 1		14837	2.1769	-7471B	PPB	05 Jul 2023 15:21:37	U	SMPL
O3510-16DLX5 - 1		30597	4.6020	-7471B	PPB	05 Jul 2023 15:23:54	U	SMPL
CCV95 - 1	CCV95	35836	5.4081	-7471B	PPB	05 Jul 2023 15:26:10	U	SMPL
CCB85 - 1	CCB85	158	-0.0818	-7471B	PPB	05 Jul 2023 15:28:28	U	SMPL