

LB126319


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

48043



A= 0.0000e+000

B= 2.0582e-004

C= 4.2270e-002

Rho= 0.9997048

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/bD
0.0	0.000	0.066	0.066	117	0.000	117	0				6
0.2	0.200	0.213	0.013	830	0.0 %	830					-6
2.5	2.500	2.362	-0.138	11273	0.0 %	11273					0
5.0	5.000	4.987	-0.013	24027	0.0 %	24027					2
7.5	7.500	7.640	0.140	36917	0.0 %	36917					-1
10.0	10.000	9.930	-0.070	48043	0.0 %	48043					

LB126319 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jul 2023 16:01:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	117	-	0.0000	7470A	PPB	05 Jul 2023 16:08:19	S	Std
0.2 - 1	50.2	830	-	0.2000	7470A	PPB	05 Jul 2023 16:10:36	S	Std
2.5 - 1	52.5	11273	-	2.5000	7470A	PPB	05 Jul 2023 16:12:53	S	Std
5.0 - 1	55	24027	-	5.0000	7470A	PPB	05 Jul 2023 16:17:46	S	Std
7.5 - 1	57.5	36917	-	7.5000	7470A	PPB	05 Jul 2023 16:20:02	S	Std
10.0 - 1	510	48043	-	10.0000	7470A	PPB	05 Jul 2023 16:25:02	S	Std
ICV56 - 1	ICV56	19324	4.0195	-	7470A	PPB	05 Jul 2023 16:32:56	U	SMPL
ICB56 - 1	ICB56	-35	0.0351	-	7470A	PPB	05 Jul 2023 16:35:11	U	SMPL
CCV86 - 1	CCV86	23824	4.9456	-	7470A	PPB	05 Jul 2023 16:37:26	U	SMPL
CCB86 - 1	CCB86	-26	0.0369	-	7470A	PPB	05 Jul 2023 16:39:41	U	SMPL
CRA - 1	CRA	783	0.2034	-	7470A	PPB	05 Jul 2023 16:41:57	U	SMPL
HIGH STD - 1	HIGH STD	47573	9.8336	-	7470A	PPB	05 Jul 2023 16:44:12	U	SMPL
CHK STD - 1	CHK STD	34287	7.0991	-	7470A	PPB	05 Jul 2023 16:46:28	U	SMPL
PB153669TB - 1	PB153669TB	-114	0.0188	-	7470A	PPB	05 Jul 2023 16:48:47	U	SMPL
CCV87 - 1	CCV87	23551	4.8895	-	7470A	PPB	05 Jul 2023 16:52:38	U	SMPL
CCB87 - 1	CCB87	-12	0.0398	-	7470A	PPB	05 Jul 2023 16:54:53	U	SMPL