



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126323

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/05/2023	17:15
2	CAL2	1	Water	NA	NA	20.7	6.99	07/05/2023	17:16
3	CAL3	1	Water	NA	NA	20.7	10.02	07/05/2023	17:18
4	ICV	1	Water	NA	NA	20.7	7.01	07/05/2023	17:20
5	CCV1	1	Water	NA	NA	20.9	2.02	07/05/2023	17:35
6	O3520-01	1	Water	NA	NA	21.6	7.09	07/05/2023	17:37
7	O3520-01DUP	1	Water	NA	NA	21.7	7.10	07/05/2023	17:38
8	CCV2	1	Water	NA	NA	20.7	12.02	07/05/2023	17:40

126323

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O3488 WorkList ID : 171617 Department : Wet-Chemistry Date : 07-05-2023 16:56:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3488-01	WC-1-COMP	Solid	pH	Cool 4 deg C	EART12	I11	06/29/2023	9041A

Date/Time 07/05/23 17:10
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC

Date/Time 07/05/23 18:20
 Raw Sample Received by: JPC
 Raw Sample Relinquished by: JPC