


LB126327

7470A

INSTRUMENT ID: CV3

Linear



A= 0.0000e+000

B= 2.3747e-004

C= -7.1775e-002

Rho= 0.9989656

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.016	-0.016	235	0.000	235	0				
0.2	0.200	0.208	0.008	1179	0.0 %	1179					
2.5	2.500	2.391	-0.109	10373	0.0 %	10373					
5.0	5.000	5.325	0.325	22727	0.0 %	22727					
7.5	7.500	7.275	-0.225	30936	0.0 %	30936					
10.0	10.000	10.017	0.017	42483	0.0 %	42483					

141730

LB126327 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Jul 2023 12:37:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	235	-	0.0000	7470A	PPB	06 Jul 2023 13:18:01	S	Std
0.2 - 1	S0-2	1179	-	0.2000	7470A	PPB	06 Jul 2023 13:22:24	S	Std
2.5 - 1	S2.5	10373	-	2.5000	7470A	PPB	06 Jul 2023 13:29:07	S	Std
5.0 - 1	S5	22727	-	5.0000	7470A	PPB	06 Jul 2023 13:31:25	S	Std
7.5 - 1	S7.5	30936	-	7.5000	7470A	PPB	06 Jul 2023 13:36:21	S	Std
10.0 - 1	S10	42483	-	10.0000	7470A	PPB	06 Jul 2023 13:46:27	S	Std
ICV58 - 1	ICV58	18367	4.2898	-7470A	PPB	06 Jul 2023 14:05:41	U	SMPL	
ICB58 - 1	ICB58	14	-0.0685	-7470A	PPB	06 Jul 2023 14:07:56	U	SMPL	
CCV92 - 1	CCV92	21160	4.9531	-7470A	PPB	06 Jul 2023 14:10:12	U	SMPL	
CCB92 - 1	CCB92	54	-0.0590	-7470A	PPB	06 Jul 2023 14:15:04	U	SMPL	
CRA - 1	CRA	1047	0.1769	-7470A	PPB	06 Jul 2023 14:22:30	U	SMPL	
HighStd - 1	HighStd	47815	11.2828	-7470A	PPB	06 Jul 2023 14:32:24	U	SMPL	
ChkStd - 1	ChkStd	32463	7.6372	-7470A	PPB	06 Jul 2023 14:44:35	U	SMPL	
PB154003BL - 1	PBW	-10	-0.0742	-7470A	PPB	06 Jul 2023 14:46:50	U	SMPL	
PB154003BS - 1	LCSW	18524	4.3271	-7470A	PPB	06 Jul 2023 14:49:07	U	SMPL	
O3507-02 - 1	CP-1	46	-0.0609	-7470A	PPB	06 Jul 2023 14:51:24	U	SMPL	
O3507-05 - 1	CP-2	64	-0.0566	-7470A	PPB	06 Jul 2023 14:53:40	U	SMPL	
O3507-08 - 1	CP-3	86	-0.0514	-7470A	PPB	06 Jul 2023 14:55:58	U	SMPL	
O3509-02 - 1	WC-WWS-01	141	-0.0383	-7470A	PPB	06 Jul 2023 14:58:15	U	SMPL	
O3509-07 - 1	WC-WWS-02	266	-0.0086	-7470A	PPB	06 Jul 2023 15:00:30	U	SMPL	
CCV93 - 1	CCV93	21958	5.1426	-7470A	PPB	06 Jul 2023 15:02:45	U	SMPL	
CCB93 - 1	CCB93	-33	-0.0796	-7470A	PPB	06 Jul 2023 15:05:00	U	SMPL	
O3509-12 - 1	WC-WWS-03	215	-0.0207	-7470A	PPB	06 Jul 2023 15:07:16	U	SMPL	
O3509-17 - 1	WC-WWS-04	261	-0.0098	-7470A	PPB	06 Jul 2023 15:09:32	U	SMPL	
O3509-22 - 1	WC-WWS-05	325	0.0054	-7470A	PPB	06 Jul 2023 15:11:46	U	SMPL	
O3510-02 - 1	WC-WWS-06	201	-0.0240	-7470A	PPB	06 Jul 2023 15:14:02	U	SMPL	
O3510-07 - 1	WC-WWS-07	324	0.0052	-7470A	PPB	06 Jul 2023 15:16:18	U	SMPL	
O3510-12 - 1	WC-WWS-10	305	0.0007	-7470A	PPB	06 Jul 2023 15:18:34	U	SMPL	
O3510-17 - 1	WC-WWS-11	250	-0.0124	-7470A	PPB	06 Jul 2023 15:20:50	U	SMPL	
O3514-02 - 1	CP-4	141	-0.0383	-7470A	PPB	06 Jul 2023 15:23:07	U	SMPL	
O3514-05 - 1	CP-5	183	-0.0283	-7470A	PPB	06 Jul 2023 15:25:24	U	SMPL	
O3514-05DUP - 1	CP-5DUP	206	-0.0229	-7470A	PPB	06 Jul 2023 15:27:39	U	SMPL	
CCV94 - 1	CCV94	20302	4.7493	-7470A	PPB	06 Jul 2023 15:32:40	U	SMPL	
CCB94 - 1	CCB94	-92	-0.0936	-7470A	PPB	06 Jul 2023 15:34:55	U	SMPL	
O3514-05MS - 1	CP-5MS	15947	3.7151	-7470A	PPB	06 Jul 2023 15:37:12	U	SMPL	
O3514-05MSD - 1	CP-5MSD	14789	3.4401	-7470A	PPB	06 Jul 2023 15:39:28	U	SMPL	
PB153981TB - 1	PB153981TB	132	-0.0404	-7470A	PPB	06 Jul 2023 15:41:44	U	SMPL	
O3514-05LX5 - 1		19	-0.0673	-7470A	PPB	06 Jul 2023 15:44:00	U	SMPL	
O3514-05A - 1		16819	3.9222	-7470A	PPB	06 Jul 2023 15:46:15	U	SMPL	
CCV95 - 1	CCV95	22763	5.3337	-7470A	PPB	06 Jul 2023 15:53:40	U	SMPL	
CCB95 - 1	CCB95	145	-0.0373	-7470A	PPB	06 Jul 2023 15:58:30	U	SMPL	