

LB126333


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

71031



A= 0.0000e+000

B= 1.4353e-004

C= -4.8415e-002

Rho= 0.9994113

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%v/v
0.0	0.000	-0.008	-0.008	282	0.000	282					
0.2	0.200	0.193	-0.007	1682	0.0 %	1682					-4
2.5	2.500	2.612	0.112	18536	0.0 %	18536					4
5.0	5.000	5.005	0.005	35210	0.0 %	35210					0
7.5	7.500	7.251	-0.249	50857	0.0 %	50857					-3
10.0	10.000	10.147	0.147	71031	0.0 %	71031					1

LB126333

Method: 7471B

Operator: Admin

Date of Analysis: 06 Jul 2023 16:16:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	282	-	0.0000	7471B	PPB	06 Jul 2023 16:20:21	S	Std
0.2 - 1	50.2	1682	-	0.2000	7471B	PPB	06 Jul 2023 16:22:38	S	Std
2.5 - 1	52.5	18536	-	2.5000	7471B	PPB	06 Jul 2023 16:24:54	S	Std
5.0 - 1	55	35210	-	5.0000	7471B	PPB	06 Jul 2023 16:27:11	S	Std
7.5 - 1	57.5	50857	-	7.5000	7471B	PPB	06 Jul 2023 16:29:28	S	Std
10.0 - 1	110	71031	-	10.0000	7471B	PPB	06 Jul 2023 16:31:46	S	Std
ICV59 - 1	ICV59	28459	4.0363	-	7471B	PPB	06 Jul 2023 16:37:21	U	SMPL
ICB59 - 1	ICB59	84	-0.0364	-	7471B	PPB	06 Jul 2023 16:39:36	U	SMPL
CCV96 - 1	CCV96	34872	4.9567	-	7471B	PPB	06 Jul 2023 16:41:52	U	SMPL
CCB96 - 1	CCB96	83	-0.0365	-	7471B	PPB	06 Jul 2023 16:44:08	U	SMPL
CRA - 1	CRA	1668	0.1910	-	7471B	PPB	06 Jul 2023 16:46:25	U	SMPL
HighStd - 1	HighStd	71871	10.2672	-	7471B	PPB	06 Jul 2023 16:48:40	U	SMPL
ChkStd - 1	ChkStd	47194	6.7253	-	7471B	PPB	06 Jul 2023 16:50:55	U	SMPL
PB154024BL - 1	PBS	-62	-0.0573	-	7471B	PPB	06 Jul 2023 16:53:13	U	SMPL
PB154024BS - 1	LCSS	24665	3.4917	-	7471B	PPB	06 Jul 2023 16:55:30	U	SMPL
O3514-01 - 1	CP-4	2444	0.3024	-	7471B	PPB	06 Jul 2023 16:57:46	U	SMPL
O3514-04 - 1	CP-5	1766	0.2051	-	7471B	PPB	06 Jul 2023 17:00:04	U	SMPL
O3515-01 - 1	LEONIA-COMP	1170	0.1195	-	7471B	PPB	06 Jul 2023 17:02:20	U	SMPL
O3516-01 - 1	BERGENFIELD-COMP	707	0.0531	-	7471B	PPB	06 Jul 2023 17:04:37	U	SMPL
O3516-01DUP - 1	BERGENFIELD-COMPDUP	768	0.0618	-	7471B	PPB	06 Jul 2023 17:06:52	U	SMPL
CCV97 - 1	CCV97	34832	4.9510	-	7471B	PPB	06 Jul 2023 17:09:09	U	SMPL
CCB97 - 1	CCB97	81	-0.0368	-	7471B	PPB	06 Jul 2023 17:11:24	U	SMPL
O3516-01MS - 1	BERGENFIELD-COMPMS	26691	3.7825	-	7471B	PPB	06 Jul 2023 17:13:41	U	SMPL
O3516-01MSD - 1	BERGENFIELD-COMPMSD	26667	3.7791	-	7471B	PPB	06 Jul 2023 17:15:56	U	SMPL
O3516-04 - 1	BERGENFIELD-COMP-2	431	0.0134	-	7471B	PPB	06 Jul 2023 17:18:13	U	SMPL
O3519-01 - 1	KULLER-COMP	7312	1.0011	-	7471B	PPB	06 Jul 2023 17:20:30	U	SMPL
O3521-01 - 1	HAWTHORNE-COMP	320	-0.0025	-	7471B	PPB	06 Jul 2023 17:22:46	U	SMPL
O3523-01 - 1	WALDWICK-COMP	1229	0.1280	-	7471B	PPB	06 Jul 2023 17:25:02	U	SMPL
O3516-01LX5 - 1		107	-0.0331	-	7471B	PPB	06 Jul 2023 17:27:19	U	SMPL
O3516-01A - 1		27371	3.8801	-	7471B	PPB	06 Jul 2023 17:29:36	U	SMPL
CCV98 - 1	CCV98	34918	4.9633	-	7471B	PPB	06 Jul 2023 17:31:52	U	SMPL
CCB98 - 1	CCB98	71	-0.0382	-	7471B	PPB	06 Jul 2023 17:34:08	U	SMPL