



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB126335

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	07/06/2023	17:00
2	CAL2	1	Water	NA	NA	21.1	6.99	07/06/2023	17:02
3	CAL3	1	Water	NA	NA	20.1	10.02	07/06/2023	17:05
4	ICV	1	Water	NA	NA	21.2	7.01	07/06/2023	17:10
5	CCV1	1	Water	NA	NA	20.9	2.01	07/06/2023	17:11
6	O3525-01	1	Water	NA	NA	28.8	6.56	07/06/2023	17:15
7	O3525-02	1	Water	NA	NA	29.0	6.62	07/06/2023	17:20
8	O3525-03	1	Water	NA	NA	29.2	6.53	07/06/2023	17:25
9	O3525-04	1	Water	NA	NA	29.3	5.59	07/06/2023	17:30
10	O3533-01	1	Water	NA	NA	28.4	7.46	07/06/2023	17:33
11	O3533-03	1	Water	NA	NA	29.2	5.13	07/06/2023	17:36
12	O3546-01	1	Water	NA	NA	29.4	6.05	07/06/2023	17:40
13	O3546-03	1	Water	NA	NA	29.6	6.03	07/06/2023	17:45
14	O3546-03DUP	1	Water	NA	NA	29.7	6.04	07/06/2023	17:46
15	CCV2	1	Water	NA	NA	20.9	12.03	07/06/2023	17:50

WB 26335

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 03346 WorkList ID : 171648 Department : Wet-Chemistry Date : 07-06-2023 16:44:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3525-01	C 33975	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C
O3525-02	C 50054	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C
O3525-03	C 50055	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C
O3525-04	A 50053	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C
O3533-01	C 7623	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C
O3533-03	C 1674	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C
O3546-01	C 1675	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C
O3546-03	C 1676	Water	pH	Cool 4 deg C	PSEG03	I11	07/06/2023	9040C

Date/Time 07/06/23 16:50
 Raw Sample Received by: J. C. (5m)
 Raw Sample Relinquished by: J. C. (5m)

Date/Time 07/06/23 18:00
 Raw Sample Received by: J. C. (5m)
 Raw Sample Relinquished by: J. C. (5m)