

LB126349


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58518



A= 0.0000e+000

B= 1.7215e-004 slope

C= -7.1215e-002 y-intercept

Rho= 0.9996456

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.00	0.000	-0.054	-0.054	101	0.000	101					5
0.05	0.050					302	0				-5
0.20	0.200	0.191	-0.009	1524	0.0 %	1524					1
2.50	2.500	2.513	0.013	15012	0.0 %	15012					4
5.00	5.000	5.187	0.187	30544	0.0 %	30544					-2
7.50	7.500	7.359	-0.141	43163	0.0 %	43163					0
10.0	10.000	10.003	0.003	58518	0.0 %	58518					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	101	-		0 SFAM01.1	PPB	7/7/2023 14:06	Std
	0.2 S01	1524	-		0.2 SFAM01.1	PPB	7/7/2023 14:08	Std
	2.5 S02	15012	-		2.5 SFAM01.1	PPB	7/7/2023 14:11	Std
	5 S03	30544	-		5 SFAM01.1	PPB	7/7/2023 14:13	Std
	7.5 S04	43163	-		7.5 SFAM01.1	PPB	7/7/2023 14:18	Std
	10 S05	58518	-		10 SFAM01.1	PPB	7/7/2023 14:25	Std
ICV060	ICV060	25000	4.2327 -		SFAM01.1	PPB	7/7/2023 14:28	SMPL
ICB060	ICB060	13	-0.069 -		SFAM01.1	PPB	7/7/2023 14:31	SMPL
CCV001	CCV001	26979	4.5733 -		SFAM01.1	PPB	7/7/2023 14:33	SMPL
CCB001	CCB001	65	-0.06 -		SFAM01.1	PPB	7/7/2023 14:38	SMPL
PB154048BL	PBS048	421	0.0013 -		SFAM01.1	PPB	7/7/2023 14:40	SMPL
O3432-01	MJVA55	1760	0.2318 -		SFAM01.1	PPB	7/7/2023 14:42	SMPL
O3432-02	MJVA57	1511	0.1889 -		SFAM01.1	PPB	7/7/2023 14:45	SMPL
O3432-03	MJVA58	1440	0.1767 -		SFAM01.1	PPB	7/7/2023 14:47	SMPL
O3432-04	MJVA59	1510	0.1887 -		SFAM01.1	PPB	7/7/2023 14:49	SMPL
O3432-05	MJVA60	2014	0.2755 -		SFAM01.1	PPB	7/7/2023 14:52	SMPL
O3432-06	MJVA60D	1993	0.2719 -		SFAM01.1	PPB	7/7/2023 14:54	SMPL
O3432-07	MJVA60S	13995	2.3381 -		SFAM01.1	PPB	7/7/2023 14:56	SMPL
O3432-08	MJVA61	1512	0.1891 -		SFAM01.1	PPB	7/7/2023 14:58	SMPL
O3432-09	MJVA62	1892	0.2545 -		SFAM01.1	PPB	7/7/2023 15:01	SMPL
O3432-10	MJVA66	1477	0.1831 -		SFAM01.1	PPB	7/7/2023 15:03	SMPL
O3432-11	MJVA68	5380	0.855 -		SFAM01.1	PPB	7/7/2023 15:05	SMPL
O3432-12	MJVA89	1720	0.2249 -		SFAM01.1	PPB	7/7/2023 15:07	SMPL
PB154049BL	PBS049	303	-0.0191 -		SFAM01.1	PPB	7/7/2023 15:10	SMPL
O3439-03	A46K7	749	0.0577 -		SFAM01.1	PPB	7/7/2023 15:12	SMPL
O3439-04	A46K7D	472	0.01 -		SFAM01.1	PPB	7/7/2023 15:14	SMPL
O3439-05	A46K7S	15129	2.5333 -		SFAM01.1	PPB	7/7/2023 15:16	SMPL
O3439-06	A4630	13077	2.1801 -		SFAM01.1	PPB	7/7/2023 15:19	SMPL
O3439-07	A4631	352	-0.0106 -		SFAM01.1	PPB	7/7/2023 15:21	SMPL
O3439-08	A46H4	4478	0.6997 -		SFAM01.1	PPB	7/7/2023 15:23	SMPL
CCV002	CCV002	30546	5.1874 -		SFAM01.1	PPB	7/7/2023 15:26	SMPL
CCB002	CCB002	18	-0.0681 -		SFAM01.1	PPB	7/7/2023 15:28	SMPL
O3439-09	A46H5	2932	0.4335 -		SFAM01.1	PPB	7/7/2023 15:30	SMPL
O3439-10	A46H6	567	0.0264 -		SFAM01.1	PPB	7/7/2023 15:32	SMPL
O3439-11	A46H7	15288	2.5607 -		SFAM01.1	PPB	7/7/2023 15:35	SMPL
O3439-12	A46H8	1212	0.1374 -		SFAM01.1	PPB	7/7/2023 15:37	SMPL
O3439-13	A46G6	9954	1.6424 -		SFAM01.1	PPB	7/7/2023 15:39	SMPL
O3439-14	A46G7	10751	1.7796 -		SFAM01.1	PPB	7/7/2023 15:41	SMPL
O3439-15	A46G8	567	0.0264 -		SFAM01.1	PPB	7/7/2023 15:44	SMPL
O3439-16	A46G9	985	0.0984 -		SFAM01.1	PPB	7/7/2023 15:46	SMPL
O3439-17	A46H1	788	0.0644 -		SFAM01.1	PPB	7/7/2023 15:48	SMPL
O3439-18	A46H2	1251	0.1442 -		SFAM01.1	PPB	7/7/2023 15:50	SMPL
O3439-19	A46K9	1165	0.1293 -		SFAM01.1	PPB	7/7/2023 15:53	SMPL
O3439-20	A46L0	494	0.0138 -		SFAM01.1	PPB	7/7/2023 15:55	SMPL
O3439-21	A46L7	16002	2.6836 -		SFAM01.1	PPB	7/7/2023 15:57	SMPL
PB154050BL	PBS050	194	-0.0378 -		SFAM01.1	PPB	7/7/2023 16:00	SMPL

INSTRUMENT ID : CV1

O3425-01	A4636	2078	0.2865 -	SFAM01.1	PPB	7/7/2023 16:02 SMPL
O3425-02	A4637	1258	0.1454 -	SFAM01.1	PPB	7/7/2023 16:04 SMPL
O3425-03	A4645	10909	1.8068 -	SFAM01.1	PPB	7/7/2023 16:06 SMPL
O3425-04	A4646	31861	5.4138 -	SFAM01.1	PPB	7/7/2023 16:09 SMPL
O3425-05	A4647	25721	4.3568 -	SFAM01.1	PPB	7/7/2023 16:11 SMPL
O3425-09	A46G0	34167	5.8108 -	SFAM01.1	PPB	7/7/2023 16:13 SMPL
CCV003	CCV003	29319	4.9762 -	SFAM01.1	PPB	7/7/2023 16:15 SMPL
CCB003	CCB003	-14	-0.0736 -	SFAM01.1	PPB	7/7/2023 16:18 SMPL
O3425-10	A46G1	19885	3.3521 -	SFAM01.1	PPB	7/7/2023 16:20 SMPL
O3425-11	A46G2	2452	0.3509 -	SFAM01.1	PPB	7/7/2023 16:22 SMPL
O3425-12	A46G3	976	0.0968 -	SFAM01.1	PPB	7/7/2023 16:25 SMPL
O3425-13	A46G4	240	-0.0299 -	SFAM01.1	PPB	7/7/2023 16:27 SMPL
O3425-14	A46G5	1086	0.1157 -	SFAM01.1	PPB	7/7/2023 16:29 SMPL
O3425-15	A4639	10640	1.7605 -	SFAM01.1	PPB	7/7/2023 16:31 SMPL
O3425-16	A4640	345	-0.0118 -	SFAM01.1	PPB	7/7/2023 16:34 SMPL
O3425-17	A4641	198	-0.0371 -	SFAM01.1	PPB	7/7/2023 16:36 SMPL
O3425-18	A46E5	19095	3.2161 -	SFAM01.1	PPB	7/7/2023 16:38 SMPL
O3425-19	A46E6	5912	0.9466 -	SFAM01.1	PPB	7/7/2023 16:40 SMPL
O3425-20	A46E8	30317	5.148 -	SFAM01.1	PPB	7/7/2023 16:43 SMPL
O3425-21	A4634	9911	1.635 -	SFAM01.1	PPB	7/7/2023 16:45 SMPL
O3425-22	A4634D	10321	1.7056 -	SFAM01.1	PPB	7/7/2023 16:50 SMPL
O3425-23	A4634S	23228	3.9276 -	SFAM01.1	PPB	7/7/2023 16:52 SMPL
O3425-24	A46F2	24775	4.1939 -	SFAM01.1	PPB	7/7/2023 16:54 SMPL
CCV004	CCV004	29031	4.9266 -	SFAM01.1	PPB	7/7/2023 16:57 SMPL
CCB004	CCB004	9	-0.0697 -	SFAM01.1	PPB	7/7/2023 16:59 SMPL