



Analytical Summary Report

Analysis Method: SM2120 B
Parameter: Color
Run Number: LB126353

Reviewed By: Iwona
Supervisor Review By: Sohil
pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP102859
Calibration std, 5 color unit	WP102860
Calibration std, 10 color unit	WP102861
Calibration std, 15 color unit	WP102870
Calibration std, 20 color unit	WP102862
Calibration std, 25 color unit	WP102863
Calibration std, 30 color unit	WP102864
Calibration std, 40 color unit	WP102865
Calibration std, 50 color unit	WP102866
Calibration std, 60 color unit	WP102867
Calibration std, 70 color unit	WP102868
Calibration std, 100 color unit	WP102869

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB126353BL	LB126353BL	1	5.03	N/A	0.00	07/07/2023	14:00
2	O3556-01	QUARTERLY-INFLUENT	50	7.43	N/A	30.00	07/07/2023	14:05
3	O3560-01	QUARTERLY-OUTFALL-00	20	8.27	N/A	40.00	07/07/2023	14:10
4	O3560-01DUP	QUARTERLY-OUTFALL-00	20	8.26	N/A	40.00	07/07/2023	14:15

WORKLIST(Hardcopy Internal Chain)

LB126353

WorkList Name : COLOR-070723 WorkList ID : 171688 Department : Wet-Chemistry Date : 07-07-2023 12:01:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3556-01	QUARTERLY-INFLUENT	Water	Color	Cool 4 deg C	INFR01	I11	07/06/2023	SM2120 B
O3560-01	QUARTERLY-OUTFALL-002	Water	Color	Cool 4 deg C	INFR01	I11	07/06/2023	SM2120 B

Date/Time 07/07/23 13:30
 Raw Sample Received by: 12/20
 Raw Sample Relinquished by: 28/02

Date/Time 07/07/23 14:20
 Raw Sample Received by: 28/02
 Raw Sample Relinquished by: 12/20