


LB126356

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9389e-004

C= -5.0536e-002

Rho= 0.9998195

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%,1)
0.0	0.000	0.010	0.010	310	0.000	310	0				
0.2	0.200	0.167	-0.033	1124	0.0 %	1124					-17
2.5	2.500	2.615	0.115	13748	0.0 %	13748					5
5.0	5.000	4.904	-0.096	25555	0.0 %	25555					-2
7.5	7.500	7.447	-0.053	38670	0.0 %	38670					-1
10.0	10.000	10.056	0.056	52126	0.0 %	52126					1

LB126356 INSTRUMENT ID : CV:1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jul 2023 17:43:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	310	-	0.0000	7470A	PPB	07 Jul 2023 17:51:48	S	Std
0.2 - 1	50-2	1124	-	0.2000	7470A	PPB	07 Jul 2023 17:54:04	S	Std
2.5 - 1	52.5	13748	-	2.5000	7470A	PPB	07 Jul 2023 17:56:20	S	Std
5.0 - 1	55	25555	-	5.0000	7470A	PPB	07 Jul 2023 17:58:37	S	Std
7.5 - 1	57.5	38670	-	7.5000	7470A	PPB	07 Jul 2023 18:05:56	S	Std
10.0 - 1	510	52126	-	10.0000	7470A	PPB	07 Jul 2023 18:10:51	S	Std
ICV61 - 1	ICV61	20176	3.8614	-	7470A	PPB	07 Jul 2023 18:24:02	U	SMPL
ICB61 - 1	ICB61	182	-0.0152	-	7470A	PPB	07 Jul 2023 18:26:17	U	SMPL
CCV05 - 1	CCV05	27553	5.2918	-	7470A	PPB	07 Jul 2023 18:31:09	U	SMPL
CCB05 - 1	CCB05	153	-0.0209	-	7470A	PPB	07 Jul 2023 18:33:23	U	SMPL
CRA - 1	CRA	1408	0.2225	-	7470A	PPB	07 Jul 2023 18:38:17	U	SMPL
HighStd - 1	HighStd	50531	9.7470	-	7470A	PPB	07 Jul 2023 18:40:31	U	SMPL
ChkStd - 1	ChkStd	35186	6.7718	-	7470A	PPB	07 Jul 2023 18:42:47	U	SMPL
PB154047BL - 1	PBW	-86	-0.0672	-	7470A	PPB	07 Jul 2023 18:45:06	U	SMPL
PB154047BS - 1	LCSW	18702	3.5756	-	7470A	PPB	07 Jul 2023 18:51:18	U	SMPL
O3518-02 - 1	COMP	645	0.0745	-	7470A	PPB	07 Jul 2023 18:53:33	U	SMPL
O3525-04 - 1	50053	436	0.0340	-	7470A	PPB	07 Jul 2023 18:55:50	U	SMPL
O3533-01 - 1	7623	1626	0.2647	-	7470A	PPB	07 Jul 2023 18:58:08	U	SMPL
O3546-01 - 1	1675	211	-0.0096	-	7470A	PPB	07 Jul 2023 19:00:24	U	SMPL
CCV06 - 1	CCV06	24844	4.7665	-	7470A	PPB	07 Jul 2023 19:02:40	U	SMPL
CCB06 - 1	CCB06	167	-0.0182	-	7470A	PPB	07 Jul 2023 19:04:55	U	SMPL
O3546-03 - 1	1676	553	0.0567	-	7470A	PPB	07 Jul 2023 19:07:11	U	SMPL
O3546-03DUP - 1	1676DUP	457	0.0381	-	7470A	PPB	07 Jul 2023 19:09:26	U	SMPL
O3546-03MS - 1	1676MS	19049	3.6429	-	7470A	PPB	07 Jul 2023 19:14:15	U	SMPL
O3546-03MSD - 1	1676MSD	20103	3.8473	-	7470A	PPB	07 Jul 2023 19:16:29	U	SMPL
O3546-03LX5 - 1		5	-0.0496	-	7470A	PPB	07 Jul 2023 19:18:46	U	SMPL
O3546-03A - 1		21922	4.2000	-	7470A	PPB	07 Jul 2023 19:21:03	U	SMPL
CCV07 - 1	CCV07	24237	4.6488	-	7470A	PPB	07 Jul 2023 19:23:19	U	SMPL
CCB07 - 1	CCB07	124	-0.0265	-	7470A	PPB	07 Jul 2023 19:25:36	U	SMPL