



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126357

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/07/2023	15:50
2	CAL2	1	Water	NA	NA	20.8	7.00	07/07/2023	15:51
3	CAL3	1	Water	NA	NA	20.7	10.02	07/07/2023	15:52
4	ICV	1	Water	NA	NA	20.9	6.98	07/07/2023	16:00
5	CCV1	1	Water	NA	NA	20.6	2.01	07/07/2023	16:01
6	O3541-02	1	Solid	20.02	20	28.0	7.26	07/07/2023	16:10
7	O3541-07	1	Solid	20.03	20	27.6	7.67	07/07/2023	16:11
8	O3555-02	1	Solid	20.04	20	28.3	8.58	07/07/2023	16:15
9	O3555-04	1	Solid	20.01	20	28.4	8.36	07/07/2023	16:20
10	O3555-06	1	Solid	20.03	20	28.0	8.66	07/07/2023	16:22
11	O3555-08	1	Solid	20.04	20	28.1	8.56	07/07/2023	16:25
12	O3555-10	1	Solid	20.02	20	27.6	8.57	07/07/2023	16:30
13	O3555-12	1	Solid	20.03	20	27.8	8.58	07/07/2023	16:33
14	O3555-12DUP	1	Solid	20.04	20	27.9	8.59	07/07/2023	16:34
15	CCV2	1	Water	NA	NA	20.9	12.03	07/07/2023	16:35

W 126357

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o3541

WorkList ID : 171694

Department : Wet-Chemistry

Date : 07-07-2023 13:43:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3541-02	WC-WWS-08	Solid	Corrosivity	Cool 4 deg C	ENTA05	I11	07/06/2023	9045D
O3541-07	WC-WWS-09	Solid	Corrosivity	Cool 4 deg C	ENTA05		07/06/2023	9045D
O3555-02	ETGI-330	Solid	Corrosivity	Cool 4 deg C	PSEG03	I11	07/07/2023	9045D
O3555-04	ETGI-347	Solid	Corrosivity	Cool 4 deg C	PSEG03	I11	07/07/2023	9045D
O3555-06	ETGI-369	Solid	Corrosivity	Cool 4 deg C	PSEG03	I11	07/07/2023	9045D
O3555-08	VNJ-203	Solid	Corrosivity	Cool 4 deg C	PSEG03	I11	07/07/2023	9045D
O3555-10	ETGI-289	Solid	Corrosivity	Cool 4 deg C	PSEG03	I11	07/07/2023	9045D
O3555-12	VNJ-240	Solid	Corrosivity	Cool 4 deg C	PSEG03	I11	07/07/2023	9045D

Date/Time 07/07/23 15:40

Raw Sample Received by: 90/690

Raw Sample Relinquished by: CP SM

Date/Time 07/07/23 17:10

Raw Sample Received by: CP SM

Raw Sample Relinquished by: 90/690