



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB126358

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/07/2023	15:50
2	CAL2	1	Water	NA	NA	20.8	7.00	07/07/2023	15:51
3	CAL3	1	Water	NA	NA	20.7	10.02	07/07/2023	15:52
4	ICV	1	Water	NA	NA	20.9	6.98	07/07/2023	16:00
5	CCV1	1	Water	NA	NA	21.1	2.02	07/07/2023	16:40
6	O3549-01	1	Water	NA	NA	27.7	4.56	07/07/2023	16:45
7	O3549-01DUP	1	Water	NA	NA	27.8	4.57	07/07/2023	16:46
8	O3549-02	1	Water	NA	NA	27.6	6.13	07/07/2023	16:47
9	O3554-01	1	Water	NA	NA	28.7	4.70	07/07/2023	16:50
10	O3562-06	1	Water	NA	NA	27.2	4.95	07/07/2023	17:00
11	O3563-01	1	Water	NA	NA	27.3	6.83	07/07/2023	17:02
12	CCV2	1	Water	NA	NA	20.9	12.02	07/07/2023	17:05

WB 126358

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH Q3549

WorkList ID : 171698

Department : Wet-Chemistry

Date : 07-07-2023 16:06:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3549-01	A 60101	Water	pH	Cool 4 deg C	PSEG03	I11	07/07/2023	9040C
Q3549-02	A 60102	Water	pH	Cool 4 deg C	PSEG03	I11	07/07/2023	9040C
Q3554-01	A POOL-WATER	Water	pH	Cool 4 deg C	STER06	I11	06/28/2023	9040C
Q3562-06	A 34590	Water	pH	Cool 4 deg C	PSEG03	I11	07/07/2023	9040C
Q3563-01	G HOUSE-BASEMENT	Water	pH	Cool 4 deg C	PSEG03	I11	07/07/2023	9040C

Date/Time 07/07/23 16:10

Raw Sample Received by: JB (000)

Raw Sample Relinquished by: CP SR

Date/Time 07/07/23 18:00

Raw Sample Received by: CP SR

Raw Sample Relinquished by: JB (000)