



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB126363
Analysis Date: 07/09/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 07/09/2023
Extraction IN Time: 07:45
Extraction OUT Time: 08:10
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB126363BL	LB126363BL	WATER	1.3	1000	100	3.2594	3.2594	3.01	3.2595	3.2595	0.0001	0.1
2	LB126363BS	LB126363BS	WATER	1.3	1000	100	3.0951	3.0951	3.02	3.1120	3.1120	0.0169	16.9
3	LB126363BSD	LB126363BSD	WATER	1.3	1000	100	3.0288	3.0288	3.03	3.0459	3.0459	0.0171	17.1
4	O3563-01	HOUSE-BASEMENT	WATER	1.6	1000	100	3.1047	3.1047	3.05	3.1181	3.1181	0.0134	13.4



Chemical Name	Chemical Lot #
HEXANE	W3034
pH Paper 0-14	M4909
Sodium Sulfate	EP2355
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>71 °C</u>	Dessicator Time In1 :	<u>09:36</u>
1.0000 gram Balance:	<u>1.0003</u>	(0.9950-1.0050)	In Time1:	<u>08:50</u>		
Bal Check Time:	<u>08:00</u>		Out OVEN TEMP1:	<u>71 °C</u>	Dessicator Time Out1:	<u>10:10</u>
			Out Time1:	<u>09:35</u>		

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2 :	<u>11:31</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time2:	<u>11:00</u>		
Bal Check Time:	<u>12:05</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2:	<u>12:00</u>
			Out Time2:	<u>11:30</u>		

VB126363

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH O3563

WorkList ID : 171718

Department : Wet-Chemistry

Date : 07-09-2023 07:32:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3563-01	HOUSE-BASEMENT	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	I11	07/07/2023	1664A

Date/Time 07/09/23 07:40
 Raw Sample Received by: JO (GOC)
 Raw Sample Relinquished by: CP SM

Date/Time 07/09/23 12:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JO (GOC)