



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126373

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	07/10/2023	13:40
2	O3562-01		1	11	100.00	>212.0	>212.0	07/10/2023	14:10
3	O3562-01DUP		1	11	100.00	>212.0	>212.0	07/10/2023	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126373

WorkList Name : fp-s-7-10

WorkList ID : 171726

Department : Wet-Chemistry

Date : 07-10-2023 08:02:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3562-01/A	34586	Solid	Flash Point	Cool 4 deg C	PSEG03	I11	07/07/2023	1010B

Date/Time 07/10/2023 13:35
 Raw Sample Received by: RM Cury
 Raw Sample Relinquished by: RM Cury

Date/Time 07/10/2023 15:00
 Raw Sample Received by: RM Cury
 Raw Sample Relinquished by: RM Cury