



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126374

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	07/10/2023	08:20
2	O3525-01		1	14	100.00	>212.0	>212.0	07/10/2023	08:50
3	O3525-01DUP		1	13	100.00	>212.0	>212.0	07/10/2023	09:20
4	O3525-02		1	13	100.00	>212.0	>212.0	07/10/2023	09:50
5	O3525-03		1	14	100.00	>212.0	>212.0	07/10/2023	10:20
6	O3525-04		1	14	100.00	>212.0	>212.0	07/10/2023	10:50
7	O3533-01		1	12	100.00	>212.0	>212.0	07/10/2023	11:20
8	O3533-03		1	13	100.00	>212.0	>212.0	07/10/2023	11:50
9	O3546-01		1	14	100.00	>212.0	>212.0	07/10/2023	12:20
10	O3546-03		1	12	100.00	>212.0	>212.0	07/10/2023	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66126374

WorkList Name : fp 7-7

WorkList ID : 171673

Department : Wet-Chemistry

Date : 07-07-2023 10:02:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3525-01D	33975	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B
O3525-02D	50054	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B
O3525-03D	50055	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B
O3525-04C	50053	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B
O3533-01F	7623	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B
O3533-03D	1674	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B
O3546-01F	1675	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B
O3546-03F	1676	Water	Flash Point	Cool 4 deg C	PSEG03	I11	07/06/2023	1010B

Date/Time 07/16/2023 08:10
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JAC

Date/Time 07/16/2023 13:36
 Raw Sample Received by: JAC
 Raw Sample Relinquished by: RM CWC