

LB126377


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

58092



A= 0.0000e+000

B= 1.7347e-004 slope

C= -8.6947e-002 y-intercept

Rho= 0.9997358

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	10.0
0.00	0.000	-0.077	-0.077	58	0.000	58					=
0.05	0.050					302	0				-23
0.20	0.200	0.154	-0.046	1388	0.0 %	1388					7
2.50	2.500	2.672	0.172	15906	0.0 %	15906					1
5.00	5.000	5.028	0.028	29489	0.0 %	29489					1
7.50	7.500	7.432	-0.068	43348	0.0 %	43348					1
10.0	10.000	9.990	-0.010	58092	0.0 %	58092					0

LB126377

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	58	-		0 SFAM01.1	PPB	7/10/2023 16:02	Std
	0.2 S01	1388	-		0.2 SFAM01.1	PPB	7/10/2023 16:04	Std
	2.5 S02	15906	-		2.5 SFAM01.1	PPB	7/10/2023 16:07	Std
	5 S03	29489	-		5 SFAM01.1	PPB	7/10/2023 16:09	Std
	7.5 S04	43348	-		7.5 SFAM01.1	PPB	7/10/2023 16:14	Std
	10 S05	58092	-		10 SFAM01.1	PPB	7/10/2023 16:19	Std
ICV063	ICV063	23343	3.9623 -		SFAM01.1	PPB	7/10/2023 16:24	SMPL
ICB063	ICB063	-25	-0.0913 -		SFAM01.1	PPB	7/10/2023 16:26	SMPL
CCV010	CCV010	29616	5.0504 -		SFAM01.1	PPB	7/10/2023 16:29	SMPL
CCB010	CCB010	-61	-0.0975 -		SFAM01.1	PPB	7/10/2023 16:31	SMPL
PB154113BL	PBS113	94	-0.0706 -		SFAM01.1	PPB	7/10/2023 16:33	SMPL
O3433-01	MJVA80	1221	0.1249 -		SFAM01.1	PPB	7/10/2023 16:35	SMPL
O3433-02	MJVA90	2525	0.3511 -		SFAM01.1	PPB	7/10/2023 16:38	SMPL
O3433-03	MJVA91	1194	0.1202 -		SFAM01.1	PPB	7/10/2023 16:40	SMPL
O3433-04	MJVA92	1713	0.2102 -		SFAM01.1	PPB	7/10/2023 16:42	SMPL
O3433-05	MJVA92D	1760	0.2184 -		SFAM01.1	PPB	7/10/2023 16:47	SMPL
O3433-06	MJVA92S	16878	2.8408 -		SFAM01.1	PPB	7/10/2023 16:49	SMPL
PB154114BL	PBS114	11	-0.085 -		SFAM01.1	PPB	7/10/2023 16:52	SMPL
O3512-03	A46X6	68119	11.7294 -		SFAM01.1	PPB	7/10/2023 16:54	SMPL
O3512-04	A46X7	17450	2.94 -		SFAM01.1	PPB	7/10/2023 16:56	SMPL
O3512-05	A46X8	23151	3.929 -		SFAM01.1	PPB	7/10/2023 16:59	SMPL
O3512-06	A46Y0	33602	5.7419 -		SFAM01.1	PPB	7/10/2023 17:01	SMPL
O3512-07	A46Y1	3155	0.4603 -		SFAM01.1	PPB	7/10/2023 17:03	SMPL
O3512-08	A46Y2	34694	5.9313 -		SFAM01.1	PPB	7/10/2023 17:05	SMPL
O3512-09	A46Y3	54110	9.2993 -		SFAM01.1	PPB	7/10/2023 17:08	SMPL
O3512-10	A46R7	5860	0.9296 -		SFAM01.1	PPB	7/10/2023 17:10	SMPL
O3512-11	A46S1	2406	0.3304 -		SFAM01.1	PPB	7/10/2023 17:12	SMPL
O3512-12	A46S2	4900	0.763 -		SFAM01.1	PPB	7/10/2023 17:14	SMPL
O3512-13	A46S3	2723	0.3854 -		SFAM01.1	PPB	7/10/2023 17:17	SMPL
O3512-14	A46T4	8079	1.3145 -		SFAM01.1	PPB	7/10/2023 17:19	SMPL
CCV011	CCV011	29354	5.005 -		SFAM01.1	PPB	7/10/2023 17:21	SMPL
CCB011	CCB011	-20	-0.0904 -		SFAM01.1	PPB	7/10/2023 17:24	SMPL
O3512-15	A46T5	4453	0.6855 -		SFAM01.1	PPB	7/10/2023 17:26	SMPL
O3512-16	A46T5D	5122	0.8015 -		SFAM01.1	PPB	7/10/2023 17:31	SMPL
O3512-17	A46T5S	19657	3.3229 -		SFAM01.1	PPB	7/10/2023 17:33	SMPL
O3512-03DLX2	A46X6	34706	5.9334 -		SFAM01.1	PPB	7/10/2023 17:35	SMPL
CCV012	CCV012	29438	5.0195 -		SFAM01.1	PPB	7/10/2023 17:37	SMPL
CCB012	CCB012	-38	-0.0935 -		SFAM01.1	PPB	7/10/2023 17:40	SMPL