



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 7/12/2023

OVENTEMP IN Celsius(°C): 108

Time IN: 13:50

In Date: 07/11/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:45

Out Date: 07/12/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB126379

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3440-01	A46A7	1	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
O3440-02	A46A4	2	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
O3440-03	A46B2	3	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
O3440-04	A46Y4	4	1.15	8.82	9.97	8.78	86.5	
O3440-05	A4715	5	1.15	8.40	9.55	8.46	87.0	
O3440-06	A4718	6	1.16	8.81	9.97	6.49	60.5	
O3440-07	A4719	7	1.12	8.60	9.72	8.74	88.6	
O3440-08	A4720	8	1.16	8.50	9.66	8.56	87.1	
O3440-09	A4721	9	1.12	8.45	9.57	8.83	91.2	
O3440-10	A4722	10	1.13	8.76	9.89	9.12	91.2	
O3440-11	A4723	11	1.18	8.52	9.7	9.06	92.5	
O3440-12	A4724	12	1.19	8.61	9.8	5.27	47.4	
O3440-13	A4725	13	1.15	8.82	9.97	7.75	74.8	
O3440-14	A4726	14	1.15	8.84	9.99	5.97	54.5	
O3440-15	A4729	15	1.18	8.46	9.64	8.62	87.9	
O3440-16	A4731	16	1.13	8.71	9.84	9.00	90.4	
O3440-17	A4733	17	1.17	8.75	9.92	8.54	84.2	
O3440-18	A4736	18	1.18	8.72	9.9	7.23	69.4	
O3440-19	A4737	19	1.16	8.82	9.98	7.49	71.8	
O3440-20	A4727	20	1.14	8.82	9.96	8.78	86.6	
O3440-21	A4727D	21	1.14	8.82	9.96	8.78	86.6	
O3440-22	A4727S	22	1.14	8.82	9.96	8.78	86.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$