



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 7/12/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 17:00
In Date: 07/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 07/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126384

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3564-01	VNJ-225	1	1.15	8.50	9.65	8.69	88.7	
O3564-02	VNJ-225	2	1.12	8.38	9.5	8.64	89.7	
O3565-01	PAT-DIS-01-7-10-2023	3	1.13	8.63	9.76	8.73	88.1	
O3565-02	PAT-DIS-01-7-10-2023-E PH-2	4	1.13	8.63	9.76	8.73	88.1	
O3566-01	61R-1A	22	1.12	8.47	9.59	8.83	91.0	
O3566-04	61R-2A	23	1.18	8.61	9.79	8.7	87.3	
O3566-07	61R-3A	24	1.19	8.61	9.8	8.34	83.0	
O3566-10	61R-4A	25	1.15	8.40	9.55	8.82	91.3	
O3566-13	61R-5A	26	1.18	8.62	9.8	8.64	86.5	
O3566-16	61R-6A	27	1.16	8.65	9.81	8.84	88.8	
O3566-19	61R-7A	28	1.15	8.80	9.95	9.28	92.4	
O3566-22	61R-8A	29	1.15	8.43	9.58	8.93	92.3	
O3566-25	61R-9A	30	1.17	8.58	9.75	8.91	90.2	
O3566-28	61R-10A	31	1.15	8.50	9.65	8.18	82.7	
O3567-02	61R-11A	32	1.15	8.78	9.93	8.88	88.0	
O3567-05	61R-12A	33	1.13	8.45	9.58	8.24	84.1	
O3567-08	61R-13A	34	1.13	8.52	9.65	8.43	85.7	
O3567-11	61R-14A	35	1.19	8.80	9.99	8.39	81.8	
O3567-14	61R-15A	36	1.19	8.71	9.9	8.31	81.7	
O3567-17	61R-16A	37	1.15	8.82	9.97	8.53	83.7	
O3567-20	61R-17A	38	1.16	8.42	9.58	8.19	83.5	
O3567-23	61R-18B	39	1.16	8.78	9.94	8.31	81.4	
O3567-24	61R-18C	40	1.18	8.47	9.65	8.42	85.5	
O3567-25	61R-19B	41	1.14	8.80	9.94	9.00	89.3	
O3567-26	61R-19C	42	1.13	8.40	9.53	7.57	76.7	
O3567-27	61R-20B	43	1.15	8.36	9.51	7.98	81.7	
O3567-28	61R-20C	44	1.15	8.83	9.98	8.41	82.2	
O3567-29	61R-21B	45	1.15	8.40	9.55	8.29	85.0	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 7/12/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 17:00
In Date: 07/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 07/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126384

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3568-01	61R-21C	46	1.19	8.80	9.99	9.26	91.7	
O3568-02	61R-22B	47	1.15	8.35	9.5	8.29	85.5	
O3568-03	61R-22C	48	1.12	8.52	9.64	8.42	85.7	
O3568-04	61R-23B	49	1.11	8.76	9.87	8.67	86.3	
O3568-05	61R-23C	50	1.11	8.76	9.87	8.22	81.2	
O3568-06	61R-24A	51	1.15	8.40	9.55	9.03	93.8	
O3568-09	61R-25A	52	1.13	8.85	9.98	7.91	76.6	
O3568-12	61R-26A	53	1.12	8.60	9.72	8.73	88.5	
O3568-15	61R-27A	54	1.11	8.77	9.88	8.97	89.6	
O3568-18	61R-DUP1	55	1.16	8.78	9.94	8.89	88.0	
O3568-19	61R-DUP2	56	1.18	8.42	9.6	8.22	83.6	
O3568-20	61R-DUP3	57	1.16	8.48	9.64	8.79	90.0	
O3568-23	61R-DUP6	58	1.16	8.45	9.61	8.5	86.9	
O3568-25	61R-DUP8	59	1.15	8.82	9.97	8.42	82.4	
O3569-01	WALL	5	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3569-03	BUILDING	6	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3570-01	EO-03-071023	7	1.18	8.66	9.84	8.97	90.0	
O3570-02	EO-03-071023-EPH-2	8	1.13	8.77	9.9	9.04	90.2	
O3570-03	EO-01-071023	9	1.19	8.60	9.79	8.77	88.1	
O3570-04	EO-01-071023-EPH-2	10	1.11	8.58	9.69	8.66	88.0	
O3572-01	LOD-MDL-SOIL-01-QT3-2023	11	1.00	1.00	2.00	2.00	100.0	
O3572-02	LOQ-SOIL-02-QT3-2023	12	1.00	1.00	2.00	2.00	100.0	
O3572-03	MDL-SOIL-03-QT3-2023	13	1.00	1.00	2.00	2.00	100.0	
O3572-04	LOD-MDL-SOIL-04-QT3-2023	14	1.00	1.00	2.00	2.00	100.0	
O3572-05	LOQ-SOIL-05-QT3-2023	15	1.00	1.00	2.00	2.00	100.0	
O3572-06	MDL-SOIL-06-QT3-2023	16	1.00	1.00	2.00	2.00	100.0	
O3572-16	LLOQ-SOIL-QT3-2023	17	1.00	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 7/12/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 17:00
In Date: 07/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 07/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126384

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3573-03	MDL-SOIL-QT3-2023-05	18	1.00	1.00	2.00	2.00	100.0	
O3573-04	MDL-SOIL-QT3-2023-06	19	1.00	1.00	2.00	2.00	100.0	
O3573-05	MDL-MED-SOIL-QT3-2023-05	20	1.00	1.00	2.00	2.00	100.0	
O3573-06	MDL-MED-SOIL-QT3-2023-06	21	1.00	1.00	2.00	2.00	100.0	
O3579-01	MIXED-DEMO	60	1.16	8.49	9.65	8.5	86.5	
O3581-01	WC-1	61	1.19	8.58	9.77	8.96	90.6	
O3581-02	WC-1-EPH-2	62	1.19	8.58	9.77	8.96	90.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$